

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062218\
 Data File : VD058303.D
 Acq On : 22 Jun 2018 18:22
 Operator : VA/AP
 Sample : J3663-06
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WP021SB472-28-30-180618

Quant Time: Jun 22 23:15:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 20 05:52:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	643162	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.69	114	878254	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.68	117	754946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.91	152	398075	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.02	65	379772	58.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.92%	
35) Dibromofluoromethane	5.54	113	378064	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
50) Toluene-d8	8.70	98	844153	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	11.92	95	382838	54.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.38%	
Target Compounds						
16) Acetone	2.61	43	32486	36.57	ug/l	# 83
65) Chlorobenzene	10.72	112	20337	1.44	ug/l	99
88) 1,4-Dichlorobenzene	12.93	146	13615	1.13	ug/l	85
91) 1,2-Dichlorobenzene	13.19	146	59106	5.60	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062218\
Data File : VD058303.D
Acq On : 22 Jun 2018 18:22
Operator : VA/AP
Sample : J3663-06
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WP021SB472-28-30-180618

Quant Time: Jun 22 23:15:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 20 05:52:56 2018
Response via : Initial Calibration

