

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062218\
 Data File : VD058311.D
 Acq On : 22 Jun 2018 22:09
 Operator : VA/AP
 Sample : J3593-01RE
 Misc : 7.08g/5ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S39-9005RE

Quant Time: Jun 23 00:08:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 20 05:52:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	515541	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.69	114	715902	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.69	117	606213	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.90	152	300236	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.01	65	319421	61.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.68%	
35) Dibromofluoromethane	5.55	113	2362	0.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	0.82%	
50) Toluene-d8	8.70	98	685771	48.65	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	97.30%	
62) 4-Bromofluorobenzene	11.93	95	298657	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
Target Compounds						
16) Acetone	2.62	43	85388	119.90	ug/l	# 92
68) m/p-Xylenes	11.00	106	13777	2.06	ug/l	78
69) o-Xylene	11.40	106	10849	1.70	ug/l	95
84) 1,2,4-Trimethylbenzene	12.62	105	25937	1.70	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062218\
Data File : VD058311.D
Acq On : 22 Jun 2018 22:09
Operator : VA/AP
Sample : J3593-01RE
Misc : 7.08g/5ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S39-9005RE

Quant Time: Jun 23 00:08:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 20 05:52:56 2018
Response via : Initial Calibration

