

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062220\
 Data File : VD065844.D
 Acq On : 22 Jun 2020 15:01
 Operator : VA/SY
 Sample : VD0622SBL01
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0622SBL01

Quant Time: Jun 23 07:01:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D060520S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 05 13:07:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	163138	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	285277	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	258722	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	114780	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	95996	59.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.10%	
35) Dibromofluoromethane	7.91	113	93719	52.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.02%	
50) Toluene-d8	10.34	98	338469	50.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.74%	
62) 4-Bromofluorobenzene	12.64	95	109867	48.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.42%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062220\
Data File : VD065844.D
Acq On : 22 Jun 2020 15:01
Operator : VA/SY
Sample : VD0622SBL01
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0622SBL01

Quant Time: Jun 23 07:01:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D060520S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 05 13:07:30 2020
Response via : Initial Calibration

