

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062221\
 Data File : VD069516.D
 Acq On : 22 Jun 2021 15:16
 Operator : VA/SY
 Sample : M2773-02
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-2

Quant Time: Jun 23 04:28:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	290448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	553017	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	604292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	260457	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	247180	68.274	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.540%
35) Dibromofluoromethane	7.914	113	218211	58.372	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.740%
50) Toluene-d8	10.338	98	732517	52.240	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.480%
62) 4-Bromofluorobenzene	12.626	95	254639	55.238	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062221\
Data File : VD069516.D
Acq On : 22 Jun 2021 15:16
Operator : VA/SY
Sample : M2773-02
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-2

Quant Time: Jun 23 04:28:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

