

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062221\
 Data File : VD069521.D
 Acq On : 22 Jun 2021 17:37
 Operator : VA/SY
 Sample : M2773-07
 Misc : 5.66G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-7

Quant Time: Jun 23 04:29:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	391010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	805588	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	836053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	354842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	303060	62.181	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.360%	
35) Dibromofluoromethane	7.914	113	298294	54.777	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.560%	
50) Toluene-d8	10.338	98	1063592	52.070	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.140%	
62) 4-Bromofluorobenzene	12.626	95	361690	53.861	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.720%	
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	9849	12.006	ug/l	96
44) Trichloroethene	9.108	130	76816	13.356	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062221\
Data File : VD069521.D
Acq On : 22 Jun 2021 17:37
Operator : VA/SY
Sample : M2773-07
Misc : 5.66G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-7

Quant Time: Jun 23 04:29:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

