

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062221\
 Data File : VD069524.D
 Acq On : 22 Jun 2021 19:01
 Operator : VA/SY
 Sample : M2773-10
 Misc : 4.85G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10

Quant Time: Jun 23 04:30:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	388801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	787443	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	835806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	363701	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	296537	61.188	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	122.380%	
35) Dibromofluoromethane	7.914	113	291533	54.769	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	109.540%	
50) Toluene-d8	10.338	98	1050052	52.592	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	105.180%	
62) 4-Bromofluorobenzene	12.626	95	362072	55.161	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	110.320%	
Target Compounds						
44) Trichloroethene	9.108	130	6495	1.155	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062221\
Data File : VD069524.D
Acq On : 22 Jun 2021 19:01
Operator : VA/SY
Sample : M2773-10
Misc : 4.85G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10

Quant Time: Jun 23 04:30:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

