

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062221\
 Data File : VD069528.D
 Acq On : 22 Jun 2021 20:53
 Operator : VA/SY
 Sample : M2795-03
 Misc : 8.69G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B2-0-2

Quant Time: Jun 23 04:31:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	400122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	802458	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	835234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	347394	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	298538	59.858	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 119.720%	
35) Dibromofluoromethane	7.914	113	292945	54.004	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.000%	
50) Toluene-d8	10.338	98	1067474	52.464	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.920%	
62) 4-Bromofluorobenzene	12.626	95	370792	55.432	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.860%	
Target Compounds						
64) Tetrachloroethene	10.873	164	34744	6.624	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062221\
Data File : VD069528.D
Acq On : 22 Jun 2021 20:53
Operator : VA/SY
Sample : M2795-03
Misc : 8.69G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B2-0-2

Quant Time: Jun 23 04:31:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

