

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
 Data File : VD076537.D
 Acq On : 22 Jun 2023 14:15
 Operator : KP/SY
 Sample : 03355-11RE
 Misc : 5.05G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-6.5-7.0RE

Quant Time: Jun 23 06:36:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	38772	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	77572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	77724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	33571	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	43908	113.030	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	226.060%#
35) Dibromofluoromethane	7.805	113	35814	77.716	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	155.440%#
50) Toluene-d8	10.269	98	110520	63.238	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	126.480%
62) 4-Bromofluorobenzene	12.569	95	36120	63.682	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	127.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
Data File : VD076537.D
Acq On : 22 Jun 2023 14:15
Operator : KP/SY
Sample : 03355-11RE
Misc : 5.05G/5.0ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-6.5-7.0RE

Quant Time: Jun 23 06:36:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

