

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
 Data File : VD076542.D
 Acq On : 22 Jun 2023 16:43
 Operator : KP/SY
 Sample : 03366-01
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB37

Quant Time: Jun 23 06:36:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	47695	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	78277	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	72208	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	29628	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	33545	70.198	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.400%
35) Dibromofluoromethane	7.805	113	29920	64.342	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	128.680%
50) Toluene-d8	10.269	98	92709	52.569	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.140%
62) 4-Bromofluorobenzene	12.575	95	27331	47.752	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.500%
Target Compounds						
					Qvalue	
16) Acetone	4.022	43	18671	164.194	ug/l	93
25) 2-Butanone	7.087	43	6878	39.106	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
Data File : VD076542.D
Acq On : 22 Jun 2023 16:43
Operator : KP/SY
Sample : 03366-01
Misc : 5.02G/5.0ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB37

Quant Time: Jun 23 06:36:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

