

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
 Data File : VD076543.D
 Acq On : 22 Jun 2023 17:13
 Operator : KP/SY
 Sample : 03363-01
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP

Quant Time: Jun 23 06:37:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	8611	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	16931	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	15565	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	8571	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	7851	91.000	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	182.000%#
35) Dibromofluoromethane	7.811	113	6153	61.174	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.340%
50) Toluene-d8	10.269	98	18872	49.474	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.940%
62) 4-Bromofluorobenzene	12.575	95	6300	50.890	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
Data File : VD076543.D
Acq On : 22 Jun 2023 17:13
Operator : KP/SY
Sample : 03363-01
Misc : 5.04G/5.0ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP

Quant Time: Jun 23 06:37:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

