

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
 Data File : VD076546.D
 Acq On : 22 Jun 2023 18:42
 Operator : KP/SY
 Sample : 03365-13
 Misc : 5.08G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP4

Quant Time: Jun 23 06:37:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	105732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	200283	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	139844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	44333	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	67514	63.732	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.460%
35) Dibromofluoromethane	7.805	113	66974	56.289	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.580%
50) Toluene-d8	10.269	98	198327	43.952	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	87.900%
62) 4-Bromofluorobenzene	12.575	95	48035	32.801	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	65.600%
Target Compounds						
16) Acetone	4.022	43	5571	22.100	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
Data File : VD076546.D
Acq On : 22 Jun 2023 18:42
Operator : KP/SY
Sample : 03365-13
Misc : 5.08G/5.0ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP4

Quant Time: Jun 23 06:37:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

