

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
 Data File : VD076549.D
 Acq On : 22 Jun 2023 20:12
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 23 06:37:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	70	0.00
2 T	Dichlorodifluoromethane	0.380	0.351	7.6	72	0.00
3 P	Chloromethane	0.790	0.728	7.8	69	0.00
4 C	Vinyl Chloride	0.881	0.802	9.0#	66	0.00
5 T	Bromomethane	0.573	0.464	19.0	63	0.00
6 T	Chloroethane	0.604	0.578	4.3	71	0.00
7 T	Trichlorofluoromethane	0.725	0.743	-2.5	81	0.00
8 T	Diethyl Ether	0.273	0.271	0.7	69	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.467	0.436	6.6	72	0.00
10 T	Methyl Iodide	0.554	0.538	2.9	66	0.00
11 T	Tert butyl alcohol	0.043	0.040	7.0	68	0.00
12 CM	1,1-Dichloroethene	0.458	0.421	8.1#	68	0.00
13 T	Acrolein	0.047	0.042	10.6	66	0.00
14 T	Allyl chloride	0.837	0.732	12.5	63	0.00
15 T	Acrylonitrile	0.140	0.139	0.7	67	0.00
16 T	Acetone	0.119	0.096	19.3	52	0.00
17 T	Carbon Disulfide	1.250	1.097	12.2	64	0.00
18 T	Methyl Acetate	0.549	0.539	1.8	67	0.00
19 T	Methyl tert-butyl Ether	1.251	1.280	-2.3	70	0.00
20 T	Methylene Chloride	0.884	0.621	29.8#	61	0.00
21 T	trans-1,2-Dichloroethene	0.529	0.502	5.1	68	0.00
22 T	Diisopropyl ether	1.809	1.721	4.9	65	0.00
23 T	Vinyl Acetate	0.897	0.858	4.3	64	0.00
24 P	1,1-Dichloroethane	1.010	0.999	1.1	70	-0.01
25 T	2-Butanone	0.184	0.175	4.9	64	0.00
26 T	2,2-Dichloropropane	0.871	0.739	15.2	62	0.00
27 T	cis-1,2-Dichloroethene	0.640	0.620	3.1	67	0.00
28 T	Bromochloromethane	0.322	0.270	16.1	63	0.00
29 T	Tetrahydrofuran	0.118	0.113	4.2	65	0.00
30 C	Chloroform	1.000	1.043	-4.3#	73	0.00
31 T	Cyclohexane	0.840	0.704	16.2	67	0.00
32 T	1,1,1-Trichloroethane	0.809	0.828	-2.3	75	0.00
33 S	1,2-Dichloroethane-d4	0.501	0.541	-8.0	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	71	0.00
35 S	Dibromofluoromethane	0.297	0.314	-5.7	75	0.00
36 T	1,1-Dichloropropene	0.416	0.379	8.9	70	0.00
37 T	Ethyl Acetate	0.241	0.201	16.6	61	0.00
38 T	Carbon Tetrachloride	0.372	0.369	0.8	74	0.00
39 T	Methylcyclohexane	0.488	0.429	12.1	68	0.00
40 TM	Benzene	1.268	1.203	5.1	69	0.00
41 T	Methacrylonitrile	0.136	0.134	1.5	67	0.00
42 TM	1,2-Dichloroethane	0.343	0.363	-5.8	73	0.00
43 T	Isopropyl Acetate	0.456	0.433	5.0	67	0.00
44 TM	Trichloroethene	0.342	0.319	6.7	69	0.00
45 C	1,2-Dichloropropane	0.344	0.333	3.2#	69	0.00
46 T	Dibromomethane	0.184	0.189	-2.7	72	0.00
47 T	Bromodichloromethane	0.441	0.457	-3.6	74	0.00
48 T	Methyl methacrylate	0.209	0.186	11.0	61	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
 Data File : VD076549.D
 Acq On : 22 Jun 2023 20:12
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 23 06:37:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	73	0.00
50 S	Toluene-d8	1.126	1.109	1.5	71	0.00
51 T	4-Methyl-2-Pentanone	0.233	0.225	3.4	66	0.00
52 CM	Toluene	0.796	0.788	1.0#	72	0.00
53 T	t-1,3-Dichloropropene	0.447	0.428	4.3	65	0.00
54 T	cis-1,3-Dichloropropene	0.530	0.498	6.0	67	0.00
55 T	1,1,2-Trichloroethane	0.258	0.264	-2.3	73	0.00
56 T	Ethyl methacrylate	0.337	0.336	0.3	68	0.00
57 T	1,3-Dichloropropane	0.441	0.436	1.1	70	0.00
58 T	2-Chloroethyl Vinyl ether	0.143	0.133	7.0	68	0.00
59 T	2-Hexanone	0.162	0.156	3.7	65	0.00
60 T	Dibromochloromethane	0.300	0.315	-5.0	73	0.00
61 T	1,2-Dibromoethane	0.242	0.257	-6.2	75	0.00
62 S	4-Bromofluorobenzene	0.366	0.388	-6.0	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	72	0.00
64 T	Tetrachloroethene	0.300	0.306	-2.0	78	0.00
65 PM	Chlorobenzene	0.950	0.910	4.2	70	0.00
66 T	1,1,1,2-Tetrachloroethane	0.340	0.349	-2.6	74	0.00
67 C	Ethyl Benzene	1.683	1.619	3.8#	72	0.00
68 T	m/p-Xylenes	0.657	0.610	7.2	70	0.00
69 T	o-Xylene	0.625	0.604	3.4	72	0.00
70 T	Styrene	1.080	1.088	-0.7	75	0.00
71 P	Bromoform	0.204	0.210	-2.9	72	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
73 T	Isopropylbenzene	3.366	3.110	7.6	73	0.00
74 T	N-amyl acetate	0.986	0.876	11.2	66	0.00
75 P	1,1,2,2-Tetrachloroethane	0.695	0.666	4.2	71	0.00
76 T	1,2,3-Trichloropropane	0.496	0.444	10.5	72	0.00
77 T	Bromobenzene	0.831	0.772	7.1	73	0.00
78 T	n-propylbenzene	4.157	3.777	9.1	73	0.00
79 T	2-Chlorotoluene	2.283	2.113	7.4	73	0.00
80 T	1,3,5-Trimethylbenzene	2.749	2.602	5.3	74	0.00
81 T	trans-1,4-Dichloro-2-butene	0.226	0.189	16.4	61	0.00
82 T	4-Chlorotoluene	2.370	2.248	5.1	74	0.00
83 T	tert-Butylbenzene	2.443	2.320	5.0	77	0.00
84 T	1,2,4-Trimethylbenzene	2.770	2.640	4.7	74	0.00
85 T	sec-Butylbenzene	3.625	3.321	8.4	73	0.00
86 T	p-Isopropyltoluene	3.029	2.831	6.5	75	0.00
87 T	1,3-Dichlorobenzene	1.631	1.542	5.5	73	0.00
88 T	1,4-Dichlorobenzene	1.627	1.555	4.4	76	0.00
89 T	n-Butylbenzene	2.954	2.641	10.6	72	0.00
90 T	Hexachloroethane	0.559	0.512	8.4	72	0.00
91 T	1,2-Dichlorobenzene	1.426	1.379	3.3	75	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.105	0.104	1.0	76	0.00
93 T	1,2,4-Trichlorobenzene	0.950	0.883	7.1	73	0.00
94 T	Hexachlorobutadiene	0.462	0.420	9.1	74	0.00
95 T	Naphthalene	1.900	1.813	4.6	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062223\
Data File : VD076549.D
Acq On : 22 Jun 2023 20:12
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jun 23 06:37:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.828	0.796	3.9	74	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6