

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062318\
 Data File : VD058359.D
 Acq On : 23 Jun 2018 23:49
 Operator : VA/AP
 Sample : J3669-05
 Misc : 6.42g/5ml/MSVOA D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WP021SB484-31-33-180620

Quant Time: Jun 25 04:07:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 20 05:52:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	775276	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	6.71	114	1046329	50.00	ug/l	0.02
63) Chlorobenzene-d5	10.70	117	862280	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.91	152	444792	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	487160	62.21	ug/l	0.02
Spiked Amount 50.000			Recovery	=	124.42%	
35) Dibromofluoromethane	5.57	113	511682	60.08	ug/l	0.02
Spiked Amount 50.000			Recovery	=	120.16%	
50) Toluene-d8	8.72	98	1175585	57.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.12%	
62) 4-Bromofluorobenzene	11.94	95	525396	62.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.57	62	116676	28.73	ug/l	96
12) 1,1-Dichloroethene	2.57	96	15862	5.96	ug/l	93
16) Acetone	2.63	43	25087	23.43	ug/l	93
27) cis-1,2-Dichloroethene	4.82	96	1323562	181.95	ug/l	90
30) Chloroform	5.35	83	14618	1.04	ug/l	93
44) Trichloroethene	7.00	130	2182229	294.74	ug/l	94
65) Chlorobenzene	10.73	112	812617	50.38	ug/l	99
87) 1,3-Dichlorobenzene	12.84	146	42771	3.09	ug/l	97
88) 1,4-Dichlorobenzene	12.93	146	105671	7.81	ug/l	99
91) 1,2-Dichlorobenzene	13.20	146	121078	10.27	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062318\
Data File : VD058359.D
Acq On : 23 Jun 2018 23:49
Operator : VA/AP
Sample : J3669-05
Misc : 6.42g/5ml/MSVOA D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WP021SB484-31-33-180620

Quant Time: Jun 25 04:07:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 20 05:52:56 2018
Response via : Initial Calibration

