

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062322\
 Data File : VD073673.D
 Acq On : 23 Jun 2022 18:45
 Operator : VA/SY
 Sample : VD0623SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0623SBL01

Quant Time: Jun 24 07:38:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 24 07:36:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	235271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	447196	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	437173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	196790	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	156553	54.177	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.360%
35) Dibromofluoromethane	7.902	113	153801	44.149	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.300%
50) Toluene-d8	10.332	98	543313	43.437	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.880%
62) 4-Bromofluorobenzene	12.620	95	181454	38.678	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062322\
Data File : VD073673.D
Acq On : 23 Jun 2022 18:45
Operator : VA/SY
Sample : VD0623SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0623SBL01

Quant Time: Jun 24 07:38:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 24 07:36:47 2022
Response via : Initial Calibration

