

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062418\
 Data File : VD058375.D
 Acq On : 24 Jun 2018 16:50
 Operator : VA/AP
 Sample : J3671-17
 Misc : 6.54g/5ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WP021SB481-15-17-180621

Manual Integrations
 APPROVED

apatel
 6/25/2018 6:21:08 PM

Quant Time: Jun 25 05:11:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 20 05:52:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	696140	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	6.71	114	926521	50.00	ug/l	0.02
63) Chlorobenzene-d5	10.70	117	673199	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	12.92	152	341085	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	432756	61.55	ug/l	0.02
Spiked Amount	50.000		Recovery	=	123.10%	
35) Dibromofluoromethane	5.56	113	455152	60.36	ug/l	0.02
Spiked Amount	50.000		Recovery	=	120.72%	
50) Toluene-d8	8.71	98	1065495	58.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	116.80%	
62) 4-Bromofluorobenzene	11.93	95	420409	56.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.82%	

Target Compounds

						Qvalue
4) Vinyl Chloride	1.56	62	220755	60.53	ug/l	96
6) Chloroethane	1.89	64	51772	37.24	ug/l	95
16) Acetone	2.62	43	50289	52.30	ug/l	96
20) Methylene Chloride	3.07	84	19338	3.25	ug/l #	76
21) trans-1,2-Dichloroethene	3.41	96	9046	3.46	ug/l	87
24) 1,1-Dichloroethane	3.99	63	75129	6.31	ug/l	96
27) cis-1,2-Dichloroethene	4.82	96	1860827	284.89	ug/l	90
39) Methylcyclohexane	7.30	83	72083	8.07	ug/l	97
40) Benzene	6.03	78	451500	20.68	ug/l	99
44) Trichloroethene	6.99	130	43734	6.67	ug/l	88
52) Toluene	8.81	92	62817	5.05	ug/l	98
65) Chlorobenzene	10.74	112	9023229	716.55	ug/l	97
67) Ethyl Benzene	10.87	91	223625	10.06	ug/l	99
68) m/p-Xylenes	11.01	106	8765	1.18	ug/l	82
69) o-Xylene	11.41	106	36106	5.10	ug/l	98
73) Isopropylbenzene	11.78	105	192997	9.46	ug/l	96
78) n-propylbenzene	12.16	91	425563	16.15	ug/l	99
80) 1,3,5-Trimethylbenzene	12.33	105	69571	4.37	ug/l	100
83) tert-Butylbenzene	12.58	119	175935	9.62	ug/l	94
84) 1,2,4-Trimethylbenzene	12.63	105	145205	8.39	ug/l	100
85) sec-Butylbenzene	12.76	105	402787	18.78	ug/l	92
86) p-Isopropyltoluene	12.89	119	43590	2.51	ug/l	97
87) 1,3-Dichlorobenzene	12.85	146	412657	38.86	ug/l	97
88) 1,4-Dichlorobenzene	12.94	146	2420221	233.40	ug/l	97
89) n-Butylbenzene	13.19	91	130932m	7.17	ug/l	
91) 1,2-Dichlorobenzene	13.21	146	6287083	695.75	ug/l	96
93) 1,2,4-Trichlorobenzene	14.35	180	143522	16.61	ug/l	100
95) Naphthalene	14.52	128	88563	7.78	ug/l	99
96) 1,2,3-Trichlorobenzene	14.66	180	19893	2.63	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062418\
Data File : VD058375.D
Acq On : 24 Jun 2018 16:50
Operator : VA/AP
Sample : J3671-17
Misc : 6.54q/5ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1


Instrument :
MSVOA_D
ClientSampleId :
WP021SB481-15-17-180621

Manual Integrations
APPROVED

apatel
6/25/2018 6:21:08 PM

Quant Time: Jun 25 05:11:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061918S.M
Quant Title : SW846 8260
QLast Update : Wed Jun 20 05:52:56 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.65 min Scan# 551

Delta R.T. 0.02 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D

ClientSampleId :

WP021SB481-15-17-180621

Tot Ion:168 Resp: 696140

Ion Ratio Lower Upper

168 100


99 53.9 41.1 61.7

Manual Integrations

APPROVED




apatel

6/25/2018 6:21:08 PM



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC



#4

Vinyl Chloride

Concen: 60.53 ug/l

RT: 1.56 min Scan# 135

Delta R.T. 0.01 min

Lab File: VD058375.D


Acq: 24 Jun 2018 16:50

Tgt Ion: 62 Resp: 220755

Ion Ratio Lower Upper



62 100


64 32.6 24.4 36.6



Abundance Ion 61.95 (61.65 to 62.65): VDC

Ion 63.95 (63.65 to 64.65): VDC





#6

Chloroethane

Concen: 37.24 ug/l

RT: 1.89 min Scan# 169

Delta R.T. 0.02 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D

ClientSampleId :

WP021SB481-15-17-180621

Tot Ion: 64 Resp: 51772

Ion Ratio Lower Upper

64 100


66 31.2 27.3 40.9

Manual Integrations

APPROVED

apatel

6/25/2018 6:21:08 PM



Abundance Ion 63.95 (63.65 to 64.65): VDC

Ion 65.90 (65.60 to 66.60): VDC

1.89

25000

20000

15000

10000

5000



0

1.80

1.90

2.00

Time-->



#16

Acetone

Concen: 52.30 ug/l

RT: 2.62 min Scan# 243

Delta R.T. 0.02 min

Lab File: VD058375.D


Acq: 24 Jun 2018 16:50

Tgt Ion: 43 Resp: 50289

Ion Ratio Lower Upper

43 100

58 14.2 12.7 19.1



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 57.95 (57.65 to 58.65): VDC

2.62

15000

10000

5000

0


2.50


2.60

2.70

2.80

Time-->









#20
Methylene Chloride
Concen: 3.25 ug/l
RT: 3.07 min Scan# 289
Delta R.T. 0.02 min
Lab File: VD058375.D
Acq: 24 Jun 2018 16:50

Instrument :
MSVOA_D
ClientSampleId :
WP021SB481-15-17-180621

Tot Ion: 84 Resp: 19338
Ion Ratio Lower Upper
84 100
49 137.2 145.0 217.6#
51 48.9 42.9 64.3
86 78.7 50.1 75.1#




Manual Integrations
APPROVED


apatel
6/25/2018 6:21:08 PM



#21
trans-1,2-Dichloroethene
Concen: 3.46 ug/l
RT: 3.41 min Scan# 323
Delta R.T. 0.03 min
Lab File: VD058375.D
Acq: 24 Jun 2018 16:50

Tgt Ion: 96 Resp: 9046
Ion Ratio Lower Upper
96 100
61 160.2 143.8 215.6
98 51.3 48.0 72.0





#24

1,1-Dichloroethane

Concen: 6.31 ug/l

RT: 3.99 min Scan# 382

Delta R.T. 0.03 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D





ClientSampleId :

WP021SB481-15-17-180621

Tot Ion:	63	Resp:	75129
Ion	Ratio	Lower	Upper
63	100		
98	5.5	1.9	5.7
100	3.5	1.3	3.9

Manual Integrations
APPROVED

apatel
6/25/2018 6:21:08 PM



#27

cis-1,2-Dichloroethene

Concen: 284.89 ug/l




RT: 4.82 min Scan# 466


Delta R.T. 0.02 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Tgt Ion:	96	Resp:	1860827
Ion	Ratio	Lower	Upper
96	100		
61	151.8	0.0	338.6
98	66.1	0.0	130.2









#33
 1,2-Dichloroethane-d4
 Concen: Below ug/l
 RT: 6.03 min Scan# 589
 Delta R.T. 0.02 min
 Lab File: VD058375.D
 Acq: 24 Jun 2018 16:50

Instrument :
 MSVOA_D
 ClientSampleId :
 WP021SB481-15-17-180621

Tot Ion	Ratio	Resp	Lower	Upper
65	100	432756		
67	48.2	0.0	97.8	




Manual Integrations
 APPROVED


apatel
 6/25/2018 6:21:08 PM



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.71 min Scan# 658
 Delta R.T. 0.02 min
 Lab File: VD058375.D
 Acq: 24 Jun 2018 16:50

Tgt Ion	Ratio	Resp	Lower	Upper
114	100	926521		
63	18.8	0.0	42.6	
88	18.1	0.0	40.6	





#35

Dibromofluoromethane

Concen: 50.00 ug/l

RT: 5.56 min Scan# 542

Delta R.T. 0.02 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D





ClientSampleId :

WP021SB481-15-17-180621

Tot Ion:	113	Resp:	455152
Ion Ratio	Lower	Upper	
113	100		
111	104.6	80.6	121.0
192	23.5	17.8	26.6

Manual Integrations
APPROVED

apatel
6/25/2018 6:21:08 PM



#39

Methylcyclohexane

Concen: 8.07 ug/l




RT: 7.30 min Scan# 718


Delta R.T. 0.02 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Tgt Ion:	83	Resp:	72083
Ion Ratio	Lower	Upper	
83	100		
55	94.9	77.4	116.2
98	40.6	35.9	53.9





#40

Benzene

Concen: 20.68 ug/l

RT: 6.03 min Scan# 589

Delta R.T. 0.02 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D

ClientSampleId :

WP021SB481-15-17-180621

Tot Ion: 78 Resp: 451500

Ion Ratio Lower Upper

78 100





77 22.4 18.5 27.7

Manual Integrations

APPROVED

apatel

6/25/2018 6:21:08 PM



#44

Trichloroethene

Concen: 6.67 ug/l

RT: 6.99 min Scan# 687

Delta R.T. 0.02 min

Lab File: VD058375.D




Acq: 24 Jun 2018 16:50


Tgt Ion:130 Resp: 43734

Ion Ratio Lower Upper

130 100

95 115.4 0.0 205.6





#50

Toluene-d8

Concen: N.D. ug/l

RT: 8.71 min Scan# 862

Delta R.T. 0.01 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D

ClientSampleId :

WP021SB481-15-17-180621

Tot Ion: 98 Resp: 1065495

Ion Ratio Lower Upper

98 100


100 67.2 53.6 80.4

Manual Integrations

APPROVED




apatel

6/25/2018 6:21:08 PM



Abundance Ion 98.05 (97.75 to 98.75): VDC

Ion 100.05 (99.75 to 100.75): VDC



#52

Toluene

Concen: 5.05 ug/l

RT: 8.81 min Scan# 872

Delta R.T. 0.01 min

Lab File: VD058375.D


Acq: 24 Jun 2018 16:50

Tgt Ion: 92 Resp: 62817

Ion Ratio Lower Upper



92 100


91 164.3 133.2 199.8



Abundance Ion 92.00 (91.70 to 92.70): VDC

Ion 91.00 (90.70 to 91.70): VDC





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 11.93 min Scan# 1189

Delta R.T. 0.01 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D





ClientSampleId :

WP021SB481-15-17-180621

Tot Ion:	95	Resp:	420409
Ion Ratio	Lower	Upper	
95	100		
174	87.7	0.0	170.2
176	86.4	0.0	164.6

Manual Integrations
APPROVED

apatel
6/25/2018 6:21:08 PM



#63

Chlorobenzene-d5

Concen: 50.00 ug/l




RT: 10.70 min Scan# 1064


Delta R.T. 0.02 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Tgt Ion:	117	Resp:	673199
Ion Ratio	Lower	Upper	
117	100		
82	56.5	47.0	70.6
119	32.0	25.3	37.9









#65
Chlorobenzene
Concen: 716.55 ug/l
RT: 10.74 min Scan# 1068
Delta R.T. 0.02 min
Lab File: VD058375.D
Acq: 24 Jun 2018 16:50

Instrument :
MSVOA_D
ClientSampleId :
WP021SB481-15-17-180621

Tot Ion:112 Resp: 9023229
Ion Ratio Lower Upper
112 100
114 32.8 25.0 37.4




Manual Integrations
APPROVED


apatel
6/25/2018 6:21:08 PM



#67
Ethyl Benzene
Concen: 10.06 ug/l
RT: 10.87 min Scan# 1081
Delta R.T. 0.02 min
Lab File: VD058375.D
Acq: 24 Jun 2018 16:50

Tgt Ion: 91 Resp: 223625
Ion Ratio Lower Upper
91 100
106 27.8 22.0 33.0






#68
m/p-Xylenes
Concen: 1.18 ug/l
RT: 11.01 min Scan# 1095
Delta R.T. -0.00 min
Lab File: VD058375.D
Acq: 24 Jun 2018 16:50

Instrument :
MSVOA_D
ClientSampleId :
WP021SB481-15-17-180621




Tot Ion:106 Resp: 8765
Ion Ratio Lower Upper
106 100
91 202.0 185.0 277.6

Manual Integrations
APPROVED

apatel
6/25/2018 6:21:08 PM




Abundance Ion 106.05 (105.75 to 106.75): V
Ion 91.00 (90.70 to 91.70): VDC






#69
o-Xylene
Concen: 5.10 ug/l
RT: 11.41 min Scan# 1136
Delta R.T. 0.01 min
Lab File: VD058375.D
Acq: 24 Jun 2018 16:50

Tgt Ion:106 Resp: 36106
Ion Ratio Lower Upper
106 100
91 236.4 120.2 360.6



Abundance Ion 106.05 (105.75 to 106.75): V
Ion 91.00 (90.70 to 91.70): VDC





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.92 min Scan# 1289

Delta R.T. 0.01 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D





ClientSampleId :

WP021SB481-15-17-180621

Tot Ion	Ratio	Lower	Upper
152	100		
115	49.3	25.1	75.3
150	168.3	0.0	308.8

Manual Integrations
APPROVED

apatel
6/25/2018 6:21:08 PM



#73

Isopropylbenzene

Concen: 9.46 ug/l




RT: 11.78 min Scan# 1173


Delta R.T. 0.01 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Tgt Ion	Ratio	Lower	Upper
105	100		
120	21.3	11.6	34.8





#78

n-propylbenzene

Concen: 16.15 ug/l

RT: 12.16 min Scan# 1212

Delta R.T. 0.01 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D

ClientSampleId :

WP021SB481-15-17-180621

Tot Ion: 91 Resp: 425563

Ion Ratio Lower Upper

91 100


120 19.8 9.7 29.1

Manual Integrations

APPROVED

apatel

6/25/2018 6:21:08 PM



Abundance Ion 91.00 (90.70 to 91.70): V

Ion 120.05 (119.75 to 120.75): V

12.16



300000

200000

100000

0

Time-->



#80

1,3,5-Trimethylbenzene

Concen: 4.37 ug/l

RT: 12.33 min Scan# 1229

Delta R.T. 0.01 min

Lab File: VD058375.D


Acq: 24 Jun 2018 16:50

Tgt Ion: 105 Resp: 69571

Ion Ratio Lower Upper

105 100

120 44.7 22.5 67.5



Abundance Ion 105.05 (104.75 to 105.75): V

Ion 120.05 (119.75 to 120.75): V

12.33

60000

50000

40000


30000


20000

10000

0

Time-->





#83

tert-Butylbenzene

Concen: 9.62 ug/l

RT: 12.58 min Scan# 1255

Delta R.T. 0.01 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D





ClientSampleId :

WP021SB481-15-17-180621

Tot Ion:	119	Resp:	175935
Ion Ratio	Lower	Upper	
119	100		
91	83.6	38.9	116.6
134	24.7	11.2	33.5

Manual Integrations
APPROVED

apatel
6/25/2018 6:21:08 PM



#84

1,2,4-Trimethylbenzene

Concen: 8.39 ug/l




RT: 12.63 min Scan# 1260


Delta R.T. 0.01 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Tgt Ion:	105	Resp:	145205
Ion Ratio	Lower	Upper	
105	100		
120	41.7	20.8	62.4









#85
 sec-Butylbenzene
 Concen: 18.78 ug/l
 RT: 12.76 min Scan# 1273
 Delta R.T. 0.01 min
 Lab File: VD058375.D
 Acq: 24 Jun 2018 16:50

Instrument :
 MSVOA_D
 ClientSampleId :
 WP021SB481-15-17-180621

Tot Ion	Ratio	Resp	Lower	Upper
105	100	402787		
134	19.6	8.1	24.3	




Manual Integrations
 APPROVED


apatel
 6/25/2018 6:21:08 PM



#86
 p-Isopropyltoluene
 Concen: 2.51 ug/l
 RT: 12.89 min Scan# 1286
 Delta R.T. 0.01 min
 Lab File: VD058375.D
 Acq: 24 Jun 2018 16:50

Tgt Ion	Ratio	Resp	Lower	Upper
119	100	43590		
134	21.5	11.6	34.8	
91	28.6	15.0	45.1	









#87
 1,3-Dichlorobenzene
 Concen: 38.86 ug/l
 RT: 12.85 min Scan# 1282
 Delta R.T. 0.01 min
 Lab File: VD058375.D
 Acq: 24 Jun 2018 16:50

Instrument :
 MSVOA_D
 ClientSampleId :
 WP021SB481-15-17-180621

Tot Ion	Ratio	Lower	Upper
146	100		
111	41.4	18.4	55.0
148	65.0	32.8	98.4




Manual Integrations
 APPROVED


apatel
 6/25/2018 6:21:08 PM



#88
 1,4-Dichlorobenzene
 Concen: 233.40 ug/l
 RT: 12.94 min Scan# 1291
 Delta R.T. 0.01 min
 Lab File: VD058375.D
 Acq: 24 Jun 2018 16:50

Tgt Ion	Ratio	Lower	Upper
146	100		
111	38.0	17.2	51.5
148	65.2	33.1	99.3





#89

n-Butylbenzene

Concen: 7.17 ug/l m

RT: 13.19 min Scan# 1317

Delta R.T. -0.00 min

Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Instrument :

MSVOA_D

ClientSampleId :





WP021SB481-15-17-180621

Tot Ion: 91 Resp: 130932

Ion	Ratio	Lower	Upper
91	100		
92	49.0	29.2	87.6
134	34.1	11.3	33.8

Manual Integrations
APPROVED

apatel
6/25/2018 6:21:08 PM



#91

1,2-Dichlorobenzene

Concen: 695.75 ug/l

RT: 13.21 min Scan# 1319




Delta R.T. 0.02 min


Lab File: VD058375.D

Acq: 24 Jun 2018 16:50

Tgt Ion: 146 Resp: 6287083

Ion	Ratio	Lower	Upper
146	100		
111	35.3	19.6	58.7
148	66.0	32.0	96.2









#93
 1,2,4-Trichlorobenzene
 Concen: 16.61 ug/l
 RT: 14.35 min Scan# 1434
 Delta R.T. 0.01 min
 Lab File: VD058375.D
 Acq: 24 Jun 2018 16:50

Instrument :
 MSVOA_D
 ClientSampleId :
 WP021SB481-15-17-180621

Tot Ion	Ratio	Lower	Upper
180	100		
182	94.2	47.1	141.3
145	22.9	11.7	35.0




Manual Integrations
 APPROVED


apatel
 6/25/2018 6:21:08 PM



#95
 Naphthalene
 Concen: 7.78 ug/l
 RT: 14.52 min Scan# 1452
 Delta R.T. 0.01 min
 Lab File: VD058375.D
 Acq: 24 Jun 2018 16:50

Tgt Ion	Ratio	Lower	Upper
128	100		
127	11.9	10.0	15.0
129	10.6	8.7	13.1





#96
 1,2,3-Trichlorobenzene
 Concen: 2.63 ug/l
 RT: 14.66 min Scan# 1466
 Delta R.T. -0.00 min
 Lab File: VD058375.D
 Acq: 24 Jun 2018 16:50

Instrument :
 MSVOA_D
 ClientSampleId :
 WP021SB481-15-17-180621

Tot Ion	Ratio	Resp Lower	Upper
180	100		
182	96.5	49.9	149.6
145	39.9	11.7	35.1

Manual Integrations
 APPROVED

apatel
 6/25/2018 6:21:08 PM

