

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062518\
 Data File : VD059410.D
 Acq On : 25 Jun 2018 20:30
 Operator : VA/AP
 Sample : J3669-03RE
 Misc : 6.45g/5ml/MSVOA D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WP021SB484-19-21-180620RE

Manual Integrations
 APPROVED

apatel
 6/29/2018 2:37:07 PM

Quant Time: Jun 28 16:42:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D062518S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 07:29:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	735153	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	6.67	114	943798	50.00	ug/l	-0.03
63) Chlorobenzene-d5	10.68	117	823847	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.90	152	447464	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.00	65	308268	41.00	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	82.00%	
35) Dibromofluoromethane	5.53	113	312147	40.60	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	81.20%	
50) Toluene-d8	8.69	98	687742	38.27	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	76.54%	
62) 4-Bromofluorobenzene	11.92	95	327338	43.79	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	87.58%	
Target Compounds						
16) Acetone	2.59	43	25551m	26.93	ug/l	Qvalue
20) Methylene Chloride	3.05	84	28464	8.76	ug/l	93
44) Trichloroethene	6.97	130	668004	99.30	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062518\
Data File : VD059410.D
Acq On : 25 Jun 2018 20:30
Operator : VA/AP
Sample : J3669-03RE
Misc : 6.45g/5ml/MSVOA D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WP021SB484-19-21-180620RE

Manual Integrations
APPROVED

apatel
6/29/2018 2:37:07 PM

Quant Time: Jun 28 16:42:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D062518S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 07:29:35 2018
Response via : Initial Calibration

