

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062519\
 Data File : VD062871.D
 Acq On : 25 Jun 2019 9:55
 Operator : VA/AP
 Sample : VD0625SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0625SBL01

Quant Time: Jun 26 05:49:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 01:50:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	931470	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.23	114	1294812	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	1041782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.53	152	546928	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	516960	51.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.12%	
35) Dibromofluoromethane	6.93	113	540968	49.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.10%	
50) Toluene-d8	10.42	98	1253717	52.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.52%	
62) 4-Bromofluorobenzene	13.56	95	541090	50.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.16%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062519\
Data File : VD062871.D
Acq On : 25 Jun 2019 9:55
Operator : VA/AP
Sample : VD0625SBL01
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0625SBL01

Quant Time: Jun 26 05:49:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 01:50:56 2019
Response via : Initial Calibration

