

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062519\
 Data File : VD062884.D
 Acq On : 25 Jun 2019 17:31
 Operator : VA/AP
 Sample : K3164-05RE
 Misc : 6.11g/5ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BAT-B07RE

Quant Time: Jun 26 06:48:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 01:50:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	686346	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.23	114	942167	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	540856	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.53	152	112482	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	415361	55.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.34%	
35) Dibromofluoromethane	6.94	113	417761	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
50) Toluene-d8	10.42	98	699476	40.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.90%	
62) 4-Bromofluorobenzene	13.56	95	179279	22.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	45.60%	
Target Compounds						
16) Acetone	3.23	43	119528	74.082	ug/l	95
44) Trichloroethene	8.55	130	85691	12.209	ug/l	96
95) Naphthalene	16.12	128	4475	1.374	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062519\
Data File : VD062884.D
Acq On : 25 Jun 2019 17:31
Operator : VA/AP
Sample : K3164-05RE
Misc : 6.11g/5ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BAT-B07RE

Quant Time: Jun 26 06:48:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 01:50:56 2019
Response via : Initial Calibration

