

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062519\
 Data File : VD062886.D
 Acq On : 25 Jun 2019 18:25
 Operator : VA/AP
 Sample : K3510-01
 Misc : 3.48g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-04-062419

Quant Time: Jun 26 06:59:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 01:50:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	123977	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.23	114	205333	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	193338	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.53	152	94160	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	148036	109.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	219.70%	
35) Dibromofluoromethane	6.93	113	106322	61.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.82%	
50) Toluene-d8	10.42	98	184228	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
62) 4-Bromofluorobenzene	13.56	95	94271	55.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.04%	
Target Compounds						
16) Acetone	3.23	43	13396	43.233	ug/l	99
20) Methylene Chloride	3.82	84	9444	9.189	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062519\
Data File : VD062886.D
Acq On : 25 Jun 2019 18:25
Operator : VA/AP
Sample : K3510-01
Misc : 3.48g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-04-062419

Quant Time: Jun 26 06:59:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 01:50:56 2019
Response via : Initial Calibration

