


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD062519\
 Data File : VD062869.D
 Acq On : 25 Jun 2019 8:44
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
 Title : SW846 8260
 Last Update : Thu Jun 20 01:50:56 2019



AutoFind: Scans 1232, 1233, 1234; Background Corrected with Scan 1227

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.2	59128	PASS
75	95	30	60	48.6	123645	PASS
95	95	100	100	100.0	254442	PASS
96	95	5	9	6.2	15827	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	85.6	217869	PASS
175	174	5	9	7.4	16221	PASS
176	174	95	101	100.7	219442	PASS
177	176	5	9	6.3	13824	PASS