

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
 Data File : VD069563.D
 Acq On : 25 Jun 2021 13:57
 Operator : VA/SY
 Sample : M2856-05
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-03-9.5-10.0

Quant Time: Jun 26 03:38:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	609479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	1182721	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	1218603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	522955	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	394394	51.914	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.820%
35) Dibromofluoromethane	7.914	113	396225	49.559	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.120%
50) Toluene-d8	10.337	98	1530358	51.031	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.060%
62) 4-Bromofluorobenzene	12.626	95	532107	53.972	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
Data File : VD069563.D
Acq On : 25 Jun 2021 13:57
Operator : VA/SY
Sample : M2856-05
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-03-9.5-10.0

Quant Time: Jun 26 03:38:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

