

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
 Data File : VD069565.D
 Acq On : 25 Jun 2021 14:54
 Operator : VA/SY
 Sample : M2856-09
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-05-9.5-10.0

Quant Time: Jun 26 03:38:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	782	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	364	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.632	117	333	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.57
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	105	10.772	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	21.540%#
35) Dibromofluoromethane	7.920	113	144	58.523	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	117.040%
50) Toluene-d8	10.344	98	531	57.533	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.060%
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
Data File : VD069565.D
Acq On : 25 Jun 2021 14:54
Operator : VA/SY
Sample : M2856-09
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-05-9.5-10.0

Quant Time: Jun 26 03:38:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

