

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
 Data File : VD069566.D
 Acq On : 25 Jun 2021 15:22
 Operator : VA/SY
 Sample : M2856-11
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-9.5-10.0

Quant Time: Jun 26 03:39:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	548088	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	1111582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	1162648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	505120	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	387435	56.710	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.420%
35) Dibromofluoromethane	7.914	113	384195	51.130	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.260%
50) Toluene-d8	10.338	98	1462704	51.897	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.800%
62) 4-Bromofluorobenzene	12.626	95	514648	55.542	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
Data File : VD069566.D
Acq On : 25 Jun 2021 15:22
Operator : VA/SY
Sample : M2856-11
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-9.5-10.0

Quant Time: Jun 26 03:39:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

