

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
 Data File : VD069568.D
 Acq On : 25 Jun 2021 16:18
 Operator : VA/SY
 Sample : M2856-15
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-08-9.5-10.0

Quant Time: Jun 26 03:39:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	912	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	2808	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	1892	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.567	152	155	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	1142	100.458	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	200.920%#
35) Dibromofluoromethane	7.920	113	1312	69.120	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	138.240%
50) Toluene-d8	10.338	98	2859	40.155	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	80.320%
62) 4-Bromofluorobenzene	12.626	95	389	16.619	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	33.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
Data File : VD069568.D
Acq On : 25 Jun 2021 16:18
Operator : VA/SY
Sample : M2856-15
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-08-9.5-10.0

Quant Time: Jun 26 03:39:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

