

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
 Data File : VD069569.D
 Acq On : 25 Jun 2021 16:46
 Operator : VA/SY
 Sample : M2856-17
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-09-9.5-10.0

Quant Time: Jun 26 03:40:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	23287	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	40568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	49373	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	20685	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	27444	94.547	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	189.100%#
35) Dibromofluoromethane	7.908	113	14879	54.257	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.520%
50) Toluene-d8	10.338	98	54826	53.300	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.600%
62) 4-Bromofluorobenzene	12.626	95	19891	58.820	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	117.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
Data File : VD069569.D
Acq On : 25 Jun 2021 16:46
Operator : VA/SY
Sample : M2856-17
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-09-9.5-10.0

Quant Time: Jun 26 03:40:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

