

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
 Data File : VD069570.D
 Acq On : 25 Jun 2021 17:15
 Operator : VA/SY
 Sample : M2856-19
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10-9.5-10.0

Quant Time: Jun 26 03:40:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	62209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	120989	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	138348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	63301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	74925	96.624	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 193.240%#	
35) Dibromofluoromethane	7.920	113	55389	67.724	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 135.440%	
50) Toluene-d8	10.338	98	153883	50.162	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.320%	
62) 4-Bromofluorobenzene	12.626	95	57140	56.656	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 113.320%	
Target Compounds						Qvalue
16) Acetone	4.156	43	11955	91.598	ug/l	# 60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
Data File : VD069570.D
Acq On : 25 Jun 2021 17:15
Operator : VA/SY
Sample : M2856-19
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10-9.5-10.0

Quant Time: Jun 26 03:40:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

