

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
 Data File : VD069574.D
 Acq On : 25 Jun 2021 19:08
 Operator : VA/SY
 Sample : M2857-07
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-14-9.5-10.0

Quant Time: Jun 26 03:41:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	355412	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	700020	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	790424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	354536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	311866	70.396	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.800%
35) Dibromofluoromethane	7.914	113	274692	58.050	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.100%
50) Toluene-d8	10.338	98	934654	52.658	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.320%
62) 4-Bromofluorobenzene	12.626	95	345681	59.240	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.480%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	11332	15.197	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
Data File : VD069574.D
Acq On : 25 Jun 2021 19:08
Operator : VA/SY
Sample : M2857-07
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-14-9.5-10.0

Quant Time: Jun 26 03:41:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

