

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
 Data File : VD069557.D
 Acq On : 25 Jun 2021 11:03
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 26 03:35:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	0.480	0.486	-1.3	117	0.00
3 P	Chloromethane	0.702	0.663	5.6	107	0.00
4 C	Vinyl Chloride	0.782	0.749	4.2#	105	0.00
5 T	Bromomethane	0.506	0.454	10.3	104	0.00
6 T	Chloroethane	0.521	0.483	7.3	102	0.00
7 T	Trichlorofluoromethane	0.929	0.980	-5.5	116	0.00
8 T	Diethyl Ether	0.246	0.257	-4.5	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.568	0.582	-2.5	117	0.00
10 T	Methyl Iodide	0.451	0.497	-10.2	106	0.00
11 T	Tert butyl alcohol	0.044	0.024	45.5#	85	0.00
12 CM	1,1-Dichloroethene	0.494	0.526	-6.5#	115	0.00
13 T	Acrolein	0.035	0.031	11.4	89	0.00
14 T	Allyl chloride	0.692	0.804	-16.2	123	0.00
15 T	Acrylonitrile	0.103	0.107	-3.9	113	0.00
16 T	Acetone	0.105	0.124	-18.1	129	0.00
17 T	Carbon Disulfide	1.626	1.732	-6.5	114	0.00
18 T	Methyl Acetate	0.296	0.260	12.2	115	0.00
19 T	Methyl tert-butyl Ether	0.970	1.025	-5.7	111	0.00
20 T	Methylene Chloride	0.851	0.710	16.6	118	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.626	-5.0	112	0.00
22 T	Diisopropyl ether	1.395	1.615	-15.8	119	0.00
23 T	Vinyl Acetate	0.764	0.900	-17.8	113	0.00
24 P	1,1-Dichloroethane	1.105	1.147	-3.8	115	0.00
25 T	2-Butanone	0.133	0.145	-9.0	115	0.00
26 T	2,2-Dichloropropane	0.903	1.000	-10.7	122	0.00
27 T	cis-1,2-Dichloroethene	0.647	0.673	-4.0	112	0.00
28 T	Bromochloromethane	0.433	0.469	-8.3	120	0.00
29 T	Tetrahydrofuran	0.073	0.081	-11.0	114	0.00
30 C	Chloroform	1.168	1.192	-2.1#	115	0.00
31 T	Cyclohexane	0.839	0.906	-8.0	120	0.00
32 T	1,1,1-Trichloroethane	0.962	1.022	-6.2	117	0.00
33 S	1,2-Dichloroethane-d4	0.623	0.579	7.1	110	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	112	0.00
35 S	Dibromofluoromethane	0.338	0.326	3.6	111	0.00
36 T	1,1-Dichloropropene	0.491	0.543	-10.6	116	0.00
37 T	Ethyl Acetate	0.182	0.191	-4.9	110	0.00
38 T	Carbon Tetrachloride	0.486	0.518	-6.6	115	0.00
39 T	Methylcyclohexane	0.484	0.564	-16.5	117	0.00
40 TM	Benzene	1.415	1.501	-6.1	114	0.00
41 T	Methacrylonitrile	0.106	0.110	-3.8	128	0.00
42 TM	1,2-Dichloroethane	0.411	0.416	-1.2	111	0.00
43 T	Isopropyl Acetate	0.331	0.345	-4.2	110	0.00
44 TM	Trichloroethene	0.357	0.370	-3.6	111	0.00
45 C	1,2-Dichloropropane	0.363	0.386	-6.3#	114	0.00
46 T	Dibromomethane	0.190	0.191	-0.5	109	0.00
47 T	Bromodichloromethane	0.511	0.521	-2.0	112	0.00
48 T	Methyl methacrylate	0.157	0.180	-14.6	121	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
 Data File : VD069557.D
 Acq On : 25 Jun 2021 11:03
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 26 03:35:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	103	0.00
50 S	Toluene-d8	1.268	1.274	-0.5	114	0.00
51 T	4-Methyl-2-Pentanone	0.174	0.189	-8.6	112	0.00
52 CM	Toluene	0.873	0.948	-8.6#	114	0.00
53 T	t-1,3-Dichloropropene	0.442	0.463	-4.8	111	0.00
54 T	cis-1,3-Dichloropropene	0.531	0.566	-6.6	116	0.00
55 T	1,1,2-Trichloroethane	0.269	0.268	0.4	110	0.00
56 T	Ethyl methacrylate	0.285	0.309	-8.4	109	0.00
57 T	1,3-Dichloropropane	0.454	0.469	-3.3	112	0.00
58 T	2-Chloroethyl Vinyl ether	0.150	0.175	-16.7	118	0.00
59 T	2-Hexanone	0.117	0.137	-17.1	117	0.00
60 T	Dibromochloromethane	0.317	0.323	-1.9	110	0.00
61 T	1,2-Dibromoethane	0.248	0.245	1.2	105	0.00
62 S	4-Bromofluorobenzene	0.417	0.420	-0.7	112	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.314	0.326	-3.8	110	0.00
65 PM	Chlorobenzene	0.994	1.029	-3.5	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.362	0.380	-5.0	112	0.00
67 C	Ethyl Benzene	1.737	1.966	-13.2#	114	0.00
68 T	m/p-Xylenes	0.654	0.751	-14.8	113	0.00
69 T	o-Xylene	0.586	0.671	-14.5	111	0.00
70 T	Styrene	1.033	1.172	-13.5	110	0.00
71 P	Bromoform	0.183	0.177	3.3	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	3.445	3.897	-13.1	112	0.00
74 T	N-amyl acetate	0.699	0.757	-8.3	109	0.00
75 P	1,1,2,2-Tetrachloroethane	0.692	0.688	0.6	107	0.00
76 T	1,2,3-Trichloropropane	0.491	0.508	-3.5	108	0.00
77 T	Bromobenzene	0.782	0.810	-3.6	107	0.00
78 T	n-propylbenzene	4.324	5.053	-16.9	116	0.00
79 T	2-Chlorotoluene	2.470	2.767	-12.0	114	0.00
80 T	1,3,5-Trimethylbenzene	2.899	3.317	-14.4	112	0.00
81 T	trans-1,4-Dichloro-2-butene	0.197	0.201	-2.0	108	0.00
82 T	4-Chlorotoluene	2.648	2.949	-11.4	114	0.00
83 T	tert-Butylbenzene	2.361	2.767	-17.2	112	0.00
84 T	1,2,4-Trimethylbenzene	2.882	3.302	-14.6	112	0.00
85 T	sec-Butylbenzene	3.615	4.154	-14.9	113	0.00
86 T	p-Isopropyltoluene	3.043	3.544	-16.5	112	0.00
87 T	1,3-Dichlorobenzene	1.602	1.657	-3.4	108	0.00
88 T	1,4-Dichlorobenzene	1.596	1.617	-1.3	107	0.00
89 T	n-Butylbenzene	3.035	3.535	-16.5	114	0.00
90 T	Hexachloroethane	0.595	0.619	-4.0	111	0.00
91 T	1,2-Dichlorobenzene	1.386	1.414	-2.0	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.105	0.103	1.9	112	0.00
93 T	1,2,4-Trichlorobenzene	0.773	0.860	-11.3	111	0.00
94 T	Hexachlorobutadiene	0.485	0.522	-7.6	112	0.00
95 T	Naphthalene	1.368	1.493	-9.1	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062521\
Data File : VD069557.D
Acq On : 25 Jun 2021 11:03
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jun 26 03:35:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.681	0.725	-6.5	106	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6