

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062522\
 Data File : VD073710.D
 Acq On : 25 Jun 2022 15:31
 Operator : VA/SY
 Sample : N3454-10
 Misc : 6.89G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-14A-1B-GR

Quant Time: Jun 27 09:38:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 24 07:36:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	286953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	555412	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	565808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	283386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	173593	49.254	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.500%
35) Dibromofluoromethane	7.909	113	174911	40.426	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	80.860%
50) Toluene-d8	10.332	98	701681	45.168	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.340%
62) 4-Bromofluorobenzene	12.620	95	292174	50.145	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.280%
Target Compounds						
16) Acetone	4.168	43	8091	6.962	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062522\
Data File : VD073710.D
Acq On : 25 Jun 2022 15:31
Operator : VA/SY
Sample : N3454-10
Misc : 6.89G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-14A-1B-GR

Quant Time: Jun 27 09:38:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 24 07:36:47 2022
Response via : Initial Calibration

