

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062522\
 Data File : VD073715.D
 Acq On : 25 Jun 2022 17:58
 Operator : VA/SY
 Sample : N3455-11
 Misc : 4.84G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-EXC-2-2

Quant Time: Jun 27 09:40:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 24 07:36:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	321840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	607849	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	586068	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	277639	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	194041	49.088	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	98.180%		
35) Dibromofluoromethane	7.909	113	196747	41.550	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	83.100%		
50) Toluene-d8	10.332	98	747431	43.962	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	87.920%		
62) 4-Bromofluorobenzene	12.620	95	283382	44.440	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	88.880%		
Target Compounds					Qvalue	
16) Acetone	4.156	43	31227	42.871	ug/l	93
17) Carbon Disulfide	4.397	76	22172	2.657	ug/l #	89
25) 2-Butanone	7.209	43	8511	8.544	ug/l	93
39) Methylcyclohexane	9.344	83	28361	3.843	ug/l	94
40) Benzene	8.338	78	23517	1.222	ug/l	97
65) Chlorobenzene	11.661	112	38225	2.665	ug/l	99
84) 1,2,4-Trimethylbenzene	13.255	105	21040	1.102	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062522\
Data File : VD073715.D
Acq On : 25 Jun 2022 17:58
Operator : VA/SY
Sample : N3455-11
Misc : 4.84G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-EXC-2-2

Quant Time: Jun 27 09:40:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062322S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 24 07:36:47 2022
Response via : Initial Calibration

