


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD062619\
Data File : VD062906.D
Acq On : 26 Jun 2019 8:51
Operator : VA/AP
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
Title : SW846 8260
Last Update : Thu Jun 20 01:50:56 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.9	73778	PASS
75	95	30	60	48.9	150877	PASS
95	95	100	100	100.0	308458	PASS
96	95	5	9	6.4	19698	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	81.2	250605	PASS
175	174	5	9	7.7	19299	PASS
176	174	95	101	98.7	247250	PASS
177	176	5	9	6.3	15586	PASS