


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD062620\
 Data File : VD065872.D
 Acq On : 26 Jun 2020 09:32
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D062620S.M
 Title : SW846 8260
 Last Update : Fri Jun 26 19:16:07 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	5353	PASS
75	95	30	60	54.1	14361	PASS
95	95	100	100	100.0	26568	PASS
96	95	5	9	7.0	1850	PASS
173	174	0.00	2	0.7	146	PASS
174	95	50	100	81.5	21664	PASS
175	174	5	9	7.4	1611	PASS
176	174	95	101	97.2	21048	PASS
177	176	5	9	8.3	1754	PASS