

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062718\
 Data File : VD059445.D
 Acq On : 27 Jun 2018 15:30
 Operator : VA/AP
 Sample : J3629-08
 Misc : 5.72g/5ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A0C10-03

Quant Time: Jun 28 03:47:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D062518S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 07:29:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	511616	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.69	114	702381	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.69	117	448421	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.90	152	128824	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.01	65	247792	47.36	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	94.72%	
35) Dibromofluoromethane	5.55	113	250546	43.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.58%	
50) Toluene-d8	8.71	98	427985	32.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.00%	
62) 4-Bromofluorobenzene	11.93	95	137831	24.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	49.56%	
Target Compounds						
16) Acetone	2.60	43	52837	80.02	ug/l #	89
20) Methylene Chloride	3.07	84	15623	6.35	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062718\
Data File : VD059445.D
Acq On : 27 Jun 2018 15:30
Operator : VA/AP
Sample : J3629-08
Misc : 5.72g/5ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A0C10-03

Quant Time: Jun 28 03:47:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D062518S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 07:29:35 2018
Response via : Initial Calibration

