

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062718\
 Data File : VD059451.D
 Acq On : 27 Jun 2018 18:19
 Operator : VA/AP
 Sample : J3614-02RE
 Misc : 10.09g/5ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ZER-SB-01-1-2RE

Quant Time: Jun 28 04:07:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D062518S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 07:29:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	624877	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.70	114	820439	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.69	117	612510	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.91	152	248007	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.02	65	246801	38.62	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	77.24%	
35) Dibromofluoromethane	5.55	113	260939	39.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	78.10%	
50) Toluene-d8	8.70	98	517963	33.15	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	66.30%	
62) 4-Bromofluorobenzene	11.93	95	196016	30.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	60.32%	
Target Compounds						
16) Acetone	2.62	43	62688	77.73	ug/l	92
17) Carbon Disulfide	2.77	76	22063	3.50	ug/l	98
20) Methylene Chloride	3.07	84	11330	2.71	ug/l	90
25) 2-Butanone	4.84	43	17210	10.19	ug/l	93
40) Benzene	6.02	78	32376	1.81	ug/l #	87
95) Naphthalene	14.51	128	9174	1.16	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062718\
Data File : VD059451.D
Acq On : 27 Jun 2018 18:19
Operator : VA/AP
Sample : J3614-02RE
Misc : 10.09g/5ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ZER-SB-01-1-2RE

Quant Time: Jun 28 04:07:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D062518S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 07:29:35 2018
Response via : Initial Calibration

