

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062818\
 Data File : VD059478.D
 Acq On : 28 Jun 2018 22:32
 Operator : VA/AP
 Sample : J3699-01RE
 Misc : 10.87g/5ml/MSVOA D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ZER-SB-04-1-2RE

Quant Time: Jun 29 02:29:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D062518S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 07:29:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	543817	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.70	114	712366	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.69	117	495475	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.91	152	151672	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.02	65	247145	44.44	ug/l	-0.01
Spiked Amount				50.000		
			Recovery	=	88.88%	
35) Dibromofluoromethane	5.55	113	264923	45.66	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.32%	
50) Toluene-d8	8.71	98	506706	37.35	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	74.70%	
62) 4-Bromofluorobenzene	11.93	95	166623	29.53	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	59.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.60	43	62059	88.42	ug/l	# 86
17) Carbon Disulfide	2.75	76	12903	2.35	ug/l	97
20) Methylene Chloride	3.06	84	12867	4.33	ug/l	83
25) 2-Butanone	4.84	43	17470	11.88	ug/l	96
40) Benzene	6.02	78	24022	1.55	ug/l	97
95) Naphthalene	14.51	128	6570	1.36	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062818\
Data File : VD059478.D
Acq On : 28 Jun 2018 22:32
Operator : VA/AP
Sample : J3699-01RE
Misc : 10.87g/5ml/MSVOA D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ZER-SB-04-1-2RE

Quant Time: Jun 29 02:29:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D062518S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 07:29:35 2018
Response via : Initial Calibration

