

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD062819\
 Data File : VD062972.D
 Acq On : 28 Jun 2019 18:04
 Operator : VA/AP
 Sample : K3554-09
 Misc : 6.28g/5ml/MSVOA D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SAMPLE-16

Quant Time: Jun 29 01:04:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 01:50:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.02	168	564192	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	8.20	114	863102	50.00	ug/l	-0.02
63) Chlorobenzene-d5	12.35	117	746286	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.51	152	394206	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.43	65	429095	69.97	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	139.94%	
35) Dibromofluoromethane	6.90	113	428793	58.92	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	117.84%	
50) Toluene-d8	10.39	98	791326	49.96	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	99.92%	
62) 4-Bromofluorobenzene	13.54	95	373873	51.91	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	103.82%	
Target Compounds						
16) Acetone	3.21	43	34913	21.684	ug/l	Qvalue 100
20) Methylene Chloride	3.79	84	24986	5.342	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD062819\
Data File : VD062972.D
Acq On : 28 Jun 2019 18:04
Operator : VA/AP
Sample : K3554-09
Misc : 6.28g/5ml/MSVOA D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SAMPLE-16

Quant Time: Jun 29 01:04:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 01:50:56 2019
Response via : Initial Calibration

