


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD062819\
 Data File : VD062954.D
 Acq On : 28 Jun 2019 9:13
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
 Title : SW846 8260
 Last Update : Thu Jun 20 01:50:56 2019



AutoFind: Scans 1230, 1231, 1232; Background Corrected with Scan 1225

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.8	42402	PASS
75	95	30	60	47.6	92768	PASS
95	95	100	100	100.0	194856	PASS
96	95	5	9	6.3	12198	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	89.8	175005	PASS
175	174	5	9	6.5	11368	PASS
176	174	95	101	95.1	166485	PASS
177	176	5	9	5.8	9734	PASS