

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062823\
 Data File : VD076618.D
 Acq On : 28 Jun 2023 16:02
 Operator : KP/SY
 Sample : 03422-12
 Misc : 5.92G/10.0ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A46K7

Quant Time: Jun 29 01:30:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM061523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 16 01:50:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	305484	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	277153	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	130224	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	86200	11.587	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	46.360%
7) Chloroethane-d5	2.793	69	85815	14.266	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.080%
11) 1,1-Dichloroethene-d2	3.911	65	27257	14.036	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.160%
21) 2-Butanone-d5	6.981	46	65322	57.977	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.960%
24) Chloroform-d	7.563	84	154713	19.223	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.880%
26) 1,2-Dichloroethane-d4	8.228	65	86436	21.770	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.080%
32) Benzene-d6	8.199	84	288255	17.002	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	68.000%
36) 1,2-Dichloropropane-d6	9.210	67	102606	18.960	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	75.840%
41) Toluene-d8	10.269	98	261068	16.630	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	66.520%
43) trans-1,3-Dichloroprop...	10.522	79	41676	18.515	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.080%
47) 2-Hexanone-d5	10.875	63	47537	56.588	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.180%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	99587	25.866	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	103.480%
66) 1,2-Dichlorobenzene-d4	13.816	152	93320	19.551	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.200%
Target Compounds					Qvalue	
13) Acetone	4.022	43	6135	6.742	ug/L	74


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062823\
Data File : VD076618.D
Acq On : 28 Jun 2023 16:02
Operator : KP/SY
Sample : 03422-12
Misc : 5.92G/10.0ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A46K7

Quant Time: Jun 29 01:30:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM061523SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 16 01:50:34 2023
Response via : Initial Calibration





#13
 Acetone
 Concen: 6.742 ug/L
 RT: 4.022 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VD076618.D
 Acq: 28 Jun 2023 16:02

Instrument :
 MSVOA_D
 ClientSampleId :
 A46K7

Tgt Ion: 43 Resp: 6135
 Ion Ratio Lower Upper
 43 100
 58 19.1 0.0 67.4

