

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062923\
 Data File : VD076629.D
 Acq On : 29 Jun 2023 10:36
 Operator : KP/SY
 Sample : VD0629SBL01
 Misc : 5.00G/10.0ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VBLK134

Quant Time: Jun 30 01:14:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM061523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Jun 16 01:50:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	225925	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	195088	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	91785	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	131379	23.879	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	95.520%	
7) Chloroethane-d5	2.799	69	105373	23.686	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	94.760%	
11) 1,1-Dichloroethene-d2	3.917	65	38488	26.799	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	107.200%	
21) 2-Butanone-d5	6.987	46	40364	48.441	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	96.880%	
24) Chloroform-d	7.569	84	146792	24.662	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	98.640%	
26) 1,2-Dichloroethane-d4	8.228	65	76124	25.924	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	103.680%	
32) Benzene-d6	8.205	84	311954	26.140	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	104.560%	
36) 1,2-Dichloropropane-d6	9.210	67	95748	25.136	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	100.560%	
41) Toluene-d8	10.269	98	292190	26.442	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	105.760%	
43) trans-1,3-Dichloroprop...	10.528	79	40291	25.430	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	101.720%	
47) 2-Hexanone-d5	10.875	63	28633	48.423	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	96.840%	
56) 1,1,2,2-Tetrachloroeth...	12.651	84	69917	25.799	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	103.200%	
66) 1,2-Dichlorobenzene-d4	13.816	152	86965	25.849	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	103.400%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD062923\
Data File : VD076629.D
Acq On : 29 Jun 2023 10:36
Operator : KP/SY
Sample : VD0629SBL01
Misc : 5.00G/10.0ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VBLK134

Quant Time: Jun 30 01:14:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM061523SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Jun 16 01:50:34 2023
Response via : Initial Calibration

