


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD070119\
Data File : VD062979.D
Acq On : 1 Jul 2019 9:30
Operator : VA/AP
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
Title : SW846 8260
Last Update : Thu Jun 20 01:50:56 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	40954	PASS
75	95	30	60	50.8	92576	PASS
95	95	100	100	100.0	182392	PASS
96	95	5	9	6.0	10997	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.7	158173	PASS
175	174	5	9	7.2	11385	PASS
176	174	95	101	100.0	158184	PASS
177	176	5	9	6.2	9871	PASS