


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD070218\
 Data File : VD059506.D
 Acq On : 2 Jul 2018 15:57
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µ/5ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D070218S.M
 Title : SW846 8260
 Last Update : Tue Jul 03 08:56:48 2018



AutoFind: Scans 1187, 1188, 1189; Background Corrected with Scan 1183

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	26.7	49821	PASS
75	95	30	60	53.5	99912	PASS
95	95	100	100	100.0	186586	PASS
96	95	5	9	6.5	12088	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	84.6	157853	PASS
175	174	5	9	7.7	12226	PASS
176	174	95	101	100.6	158730	PASS
177	176	5	9	6.2	9827	PASS