

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD070319\
 Data File : VD063023.D
 Acq On : 3 Jul 2019 14:13
 Operator : VA/AP
 Sample : K3645-02RE
 Misc : 4.31g/5ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-0D015-03-010RE

Quant Time: Jul 04 01:57:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 01:50:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	996552	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.23	114	1330398	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	1124681	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	574871	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	441012	40.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.42%	
35) Dibromofluoromethane	6.93	113	185687	16.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	33.10%	
50) Toluene-d8	10.41	98	1076344	44.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.16%	
62) 4-Bromofluorobenzene	13.55	95	442781	39.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.76%	
Target Compounds						
16) Acetone	3.23	43	70739	25.932	ug/l	96
52) Toluene	10.51	92	59363	3.493	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD070319\
Data File : VD063023.D
Acq On : 3 Jul 2019 14:13
Operator : VA/AP
Sample : K3645-02RE
Misc : 4.31g/5ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-0D015-03-010RE

Quant Time: Jul 04 01:57:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 01:50:56 2019
Response via : Initial Calibration

