

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070621\
 Data File : VD069594.D
 Acq On : 06 Jul 2021 20:59
 Operator : VA/SY
 Sample : M2946-01
 Misc : 6.92G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1

Quant Time: Jul 07 04:56:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	35990	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	72538	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	87798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	40522	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	43597	97.183	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 194.360%#	
35) Dibromofluoromethane	7.914	113	30612	62.430	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 124.860%	
50) Toluene-d8	10.338	98	97975	53.269	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.540%	
62) 4-Bromofluorobenzene	12.626	95	37641	62.251	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 124.500%	
Target Compounds						
					Qvalue	
16) Acetone	4.168	43	5133	67.980	ug/l	100
44) Trichloroethene	9.114	130	3932	7.592	ug/l	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070621\
Data File : VD069594.D
Acq On : 06 Jul 2021 20:59
Operator : VA/SY
Sample : M2946-01
Misc : 6.92G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1

Quant Time: Jul 07 04:56:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

