

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070622\
 Data File : VD073797.D
 Acq On : 06 Jul 2022 17:47
 Operator : VA/SY
 Sample : VSTDICC010
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICC010

Quant Time: Jul 07 04:59:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 07 04:53:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	112175	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	218581	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	199904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	89174	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	16016	14.168	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	28.340%#
35) Dibromofluoromethane	7.908	113	15118	12.574	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	25.140%#
50) Toluene-d8	10.326	98	53454	12.232	ug/l	0.00
Spiked Amount	50.000	Range	78 - 12	Recovery	=	24.460%#
62) 4-Bromofluorobenzene	12.620	95	20638	11.889	ug/l	0.00
Spiked Amount	50.000	Range	50 - 14	Recovery	=	23.780%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.991	85	8721	12.567	ug/l	91
3) Chloromethane	2.203	50	12397	15.342	ug/l	95
4) Vinyl Chloride	2.338	62	14066	15.343	ug/l	97
5) Bromomethane	2.750	94	11475	17.551	ug/l	98
6) Chloroethane	2.915	64	10461	15.472	ug/l	97
7) Trichlorofluoromethane	3.262	101	21373	13.078	ug/l	87
8) Diethyl Ether	3.703	74	6298	11.540	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.073	101	13032	12.168	ug/l	97
10) Methyl Iodide	4.291	142	7729	10.259	ug/l	95
11) Tert butyl alcohol	5.173	59	32688	116.267	ug/l #	93
12) 1,1-Dichloroethene	4.050	96	10282	11.222	ug/l	89
13) Acrolein	3.915	56	3448	82.871	ug/l	90
14) Allyl chloride	4.697	41	18026	11.282	ug/l	95
15) Acrylonitrile	5.420	53	16924	56.240	ug/l	94
16) Acetone	4.156	43	17785	75.959	ug/l	99
17) Carbon Disulfide	4.397	76	18131	12.581	ug/l #	95
18) Methyl Acetate	4.720	43	9952	10.979	ug/l #	77
19) Methyl tert-butyl Ether	5.479	73	37204	12.261	ug/l	93
20) Methylene Chloride	4.950	84	40183	26.829	ug/l	92
21) trans-1,2-Dichloroethene	5.467	96	12534	12.064	ug/l	87
22) Diisopropyl ether	6.356	45	47287	11.654	ug/l #	95
23) Vinyl Acetate	6.297	43	105683	50.950	ug/l	97
24) 1,1-Dichloroethane	6.250	63	27406	12.088	ug/l	99
25) 2-Butanone	7.203	43	23141	59.308	ug/l	98
26) 2,2-Dichloropropane	7.191	77	25440	12.284	ug/l	98
27) cis-1,2-Dichloroethene	7.208	96	16674	11.777	ug/l	98
28) Bromochloromethane	7.538	49	9604	11.074	ug/l #	92
29) Tetrahydrofuran	7.556	42	14275	58.618	ug/l	97
30) Chloroform	7.697	83	29895	12.341	ug/l	90
31) Cyclohexane	7.979	56	20891	11.974	ug/l #	91
32) 1,1,1-Trichloroethane	7.891	97	24313	11.525	ug/l	92
36) 1,1-Dichloropropene	8.103	75	18972	11.467	ug/l	97
37) Ethyl Acetate	7.285	43	10569	11.845	ug/l #	91
38) Carbon Tetrachloride	8.091	117	20032	11.141	ug/l	91
39) Methylcyclohexane	9.344	83	19148	10.316	ug/l	98
40) Benzene	8.344	78	54359	11.051	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070622\
 Data File : VD073797.D
 Acq On : 06 Jul 2022 17:47
 Operator : VA/SY
 Sample : VSTDIC010
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC010

Quant Time: Jul 07 04:59:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 07 04:53:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.508	41	5148	9.185	ug/l	94
42) 1,2-Dichloroethane	8.420	62	18809	12.366	ug/l	99
43) Isopropyl Acetate	8.450	43	21384	11.245	ug/l	98
44) Trichloroethene	9.102	130	14113	10.901	ug/l	83
45) 1,2-Dichloropropane	9.379	63	14994	10.626	ug/l	97
46) Dibromomethane	9.467	93	8811	11.801	ug/l	99
47) Bromodichloromethane	9.650	83	23273	11.508	ug/l	97
48) Methyl methacrylate	9.444	41	9797	11.522	ug/l	96
49) 1,4-Dioxane	9.455	88	2102	209.013	ug/l #	85
51) 4-Methyl-2-Pentanone	10.220	43	56228	57.697	ug/l	96
52) Toluene	10.397	92	35303	10.758	ug/l	100
53) t-1,3-Dichloropropene	10.608	75	21214	10.746	ug/l	98
54) cis-1,3-Dichloropropene	10.079	75	24424	10.998	ug/l	95
55) 1,1,2-Trichloroethane	10.791	97	11546	10.402	ug/l #	91
56) Ethyl methacrylate	10.649	69	14837	10.121	ug/l	93
57) 1,3-Dichloropropane	10.932	76	21134	11.288	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.932	63	29592	50.464	ug/l	100
59) 2-Hexanone	10.973	43	35427	53.541	ug/l	96
60) Dibromochloromethane	11.126	129	14567	10.755	ug/l	99
61) 1,2-Dibromoethane	11.238	107	11373	11.148	ug/l	90
64) Tetrachloroethene	10.867	164	11235	11.015	ug/l	89
65) Chlorobenzene	11.661	112	40685	10.958	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.732	131	15419	10.976	ug/l	98
67) Ethyl Benzene	11.732	91	72638	10.696	ug/l	97
68) m/p-Xylenes	11.838	106	53881	20.820	ug/l	99
69) o-Xylene	12.173	106	26501	10.486	ug/l	93
70) Styrene	12.179	104	45040	10.346	ug/l	98
71) Bromoform	12.344	173	7614	10.276	ug/l #	97
73) Isopropylbenzene	12.467	105	70822	10.444	ug/l	97
74) N-amyl acetate	12.273	43	19824	10.648	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.714	83	15165	11.084	ug/l	99
76) 1,2,3-Trichloropropane	12.773	75	18300	21.592	ug/l #	100
77) Bromobenzene	12.743	156	15359	11.066	ug/l	97
78) n-propylbenzene	12.808	91	90001	10.622	ug/l	100
79) 2-Chlorotoluene	12.896	91	50349	10.589	ug/l	99
80) 1,3,5-Trimethylbenzene	12.943	105	60759	10.749	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.520	75	3970	8.996	ug/l	95
82) 4-Chlorotoluene	12.991	91	52555	10.615	ug/l	97
83) tert-Butylbenzene	13.208	119	52691	10.497	ug/l	97
84) 1,2,4-Trimethylbenzene	13.255	105	58596	10.508	ug/l	97
85) sec-Butylbenzene	13.385	105	80616	10.464	ug/l	99
86) p-Isopropyltoluene	13.502	119	64018	10.368	ug/l	99
87) 1,3-Dichlorobenzene	13.502	146	31060	10.684	ug/l	98
88) 1,4-Dichlorobenzene	13.579	146	31971	11.034	ug/l	96
89) n-Butylbenzene	13.826	91	66028	10.599	ug/l	99
90) Hexachloroethane	14.096	117	12743	10.323	ug/l	99
91) 1,2-Dichlorobenzene	13.873	146	27255	10.542	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.490	75	2390	10.443	ug/l	93
93) 1,2,4-Trichlorobenzene	15.143	180	16098	10.581	ug/l	99
94) Hexachlorobutadiene	15.249	225	7642	10.173	ug/l	96
95) Naphthalene	15.379	128	35035	10.310	ug/l	99
96) 1,2,3-Trichlorobenzene	15.573	180	13815	10.379	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070622\
Data File : VD073797.D
Acq On : 06 Jul 2022 17:47
Operator : VA/SY
Sample : VSTDICC010
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VSTDICC010

Quant Time: Jul 07 04:59:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
Quant Title : SW846 8260
QLast Update : Thu Jul 07 04:53:08 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070622\
 Data File : VD073797.D
 Acq On : 06 Jul 2022 17:47
 Operator : VA/SY
 Sample : VSTDIC010
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC010

Quant Time: Jul 07 04:59:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 07 04:53:08 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1028
Delta R.T. 0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:168 Resp: 112175
Ion Ratio Lower Upper
168 100
99 67.8 58.4 87.6



#2
Dichlorodifluoromethane
Concen: 12.567 ug/l
RT: 1.991 min Scan# 12
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 85 Resp: 8721
Ion Ratio Lower Upper
85 100
87 27.0 16.2 48.5









#3
Chloromethane
Concen: 15.342 ug/l
RT: 2.203 min Scan# 41
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 50 Resp: 12397
Ion Ratio Lower Upper
50 100
52 32.0 27.8 41.6



#4
Vinyl Chloride
Concen: 15.343 ug/l
RT: 2.338 min Scan# 71
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 62 Resp: 14066
Ion Ratio Lower Upper
62 100
64 27.5 23.4 35.2









#5
Bromomethane
Concen: 17.551 ug/l
RT: 2.750 min Scan# 142
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 94 Resp: 11475
Ion Ratio Lower Upper
94 100
96 95.9 78.3 117.5



#6
Chloroethane
Concen: 15.472 ug/l
RT: 2.915 min Scan# 169
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 64 Resp: 10461
Ion Ratio Lower Upper
64 100
66 27.8 23.7 35.5









#7
 Trichlorofluoromethane
 Concen: 13.078 ug/l
 RT: 3.262 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VD073797.D
 Acq: 06 Jul 2022 17:47




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC010


Tgt Ion:101 Resp: 21373
 Ion Ratio Lower Upper
 101 100
 103 58.4 55.4 83.2



#8
 Diethyl Ether
 Concen: 11.540 ug/l
 RT: 3.703 min Scan# 303
 Delta R.T. 0.000 min
 Lab File: VD073797.D
 Acq: 06 Jul 2022 17:47

Tgt Ion: 74 Resp: 6298
 Ion Ratio Lower Upper
 74 100
 45 104.0 51.1 153.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 12.168 ug/l
RT: 4.073 min Scan# 366
Delta R.T. -0.012 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:101 Resp: 13032
Ion Ratio Lower Upper
101 100
85 46.7 38.8 58.2
151 64.9 54.4 81.6



#10
Methyl Iodide
Concen: 10.259 ug/l
RT: 4.291 min Scan# 403
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:142 Resp: 7729
Ion Ratio Lower Upper
142 100
127 53.5 39.7 59.5
141 13.2 11.1 16.7









#11
Tert butyl alcohol
Concen: 116.267 ug/l
RT: 5.173 min Scan# 51
Delta R.T. -0.047 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 59 Resp: 32688
Ion Ratio Lower Upper
59 100
57 0.8 2.4 3.6#



#12
1,1-Dichloroethene
Concen: 11.222 ug/l
RT: 4.050 min Scan# 362
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 96 Resp: 10282
Ion Ratio Lower Upper
96 100
61 180.8 129.3 193.9
98 61.4 47.4 71.2









#13
Acrolein
Concen: 82.871 ug/l
RT: 3.915 min Scan# 311
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC010


Tgt Ion: 56 Resp: 3448
Ion Ratio Lower Upper
56 100
55 77.1 55.4 83.2



#14
Allyl chloride
Concen: 11.282 ug/l
RT: 4.697 min Scan# 472
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 41 Resp: 18026
Ion Ratio Lower Upper
41 100
39 78.8 59.8 89.8
76 38.2 28.4 42.6









#15
Acrylonitrile
Concen: 56.240 ug/l
RT: 5.420 min Scan# 593
Delta R.T. 0.012 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC010


Tgt Ion: 53 Resp: 16924
Ion Ratio Lower Upper
53 100
52 88.5 65.9 98.9
51 39.1 29.7 44.5



#16
Acetone
Concen: 75.959 ug/l
RT: 4.156 min Scan# 380
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 43 Resp: 17785
Ion Ratio Lower Upper
43 100
58 27.4 21.3 31.9









#17
Carbon Disulfide
Concen: 12.581 ug/l
RT: 4.397 min Scan# 41
Delta R.T. -0.012 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 76 Resp: 18131
Ion Ratio Lower Upper
76 100
78 10.2 6.8 10.2#



#18
Methyl Acetate
Concen: 10.979 ug/l
RT: 4.720 min Scan# 476
Delta R.T. 0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 43 Resp: 9952
Ion Ratio Lower Upper
43 100
74 11.5 18.3 27.5#









#19
Methyl tert-butyl Ether
Concen: 12.261 ug/l
RT: 5.479 min Scan# 604
Delta R.T. 0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 73 Resp: 37204
Ion Ratio Lower Upper
73 100
57 20.5 19.0 28.6



#20
Methylene Chloride
Concen: 26.829 ug/l
RT: 4.950 min Scan# 515
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 84 Resp: 40183
Ion Ratio Lower Upper
84 100
49 127.4 97.4 146.2
51 44.6 30.8 46.2
86 70.0 47.5 71.3









#21
trans-1,2-Dichloroethene
Concen: 12.064 ug/l
RT: 5.467 min Scan# 602
Delta R.T. 0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 96 Resp: 12534
Ion Ratio Lower Upper
96 100
61 120.4 108.2 162.2
98 66.4 45.4 68.0



#22
Diisopropyl ether
Concen: 11.654 ug/l
RT: 6.356 min Scan# 754
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 45 Resp: 47287
Ion Ratio Lower Upper
45 100
43 55.1 44.6 67.0
87 21.9 23.2 34.8#
59 12.9 9.6 14.4









#23
 Vinyl Acetate
 Concen: 50.950 ug/l
 RT: 6.297 min Scan# 744
 Delta R.T. 0.000 min
 Lab File: VD073797.D
 Acq: 06 Jul 2022 17:47




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC010


Tgt Ion: 43 Resp: 105683
 Ion Ratio Lower Upper
 43 100
 86 10.6 7.6 11.4



#24
 1,1-Dichloroethane
 Concen: 12.088 ug/l
 RT: 6.250 min Scan# 736
 Delta R.T. -0.006 min
 Lab File: VD073797.D
 Acq: 06 Jul 2022 17:47

Tgt Ion: 63 Resp: 27406
 Ion Ratio Lower Upper
 63 100
 98 7.4 3.9 11.5
 100 3.7 2.2 6.6









#25
2-Butanone
Concen: 59.308 ug/l
RT: 7.203 min Scan# 898
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC010


Tgt Ion: 43 Resp: 23141
Ion Ratio Lower Upper
43 100
72 22.5 18.8 28.2



#26
2,2-Dichloropropane
Concen: 12.284 ug/l
RT: 7.191 min Scan# 896
Delta R.T. -0.012 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 77 Resp: 25440
Ion Ratio Lower Upper
77 100
97 20.8 10.8 32.4









#27
cis-1,2-Dichloroethene
Concen: 11.777 ug/l
RT: 7.208 min Scan# 898
Delta R.T. 0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 96 Resp: 16674
Ion Ratio Lower Upper
96 100
61 150.1 0.0 303.2
98 59.0 0.0 125.2



#28
Bromochloromethane
Concen: 11.074 ug/l
RT: 7.538 min Scan# 955
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 49 Resp: 9604
Ion Ratio Lower Upper
49 100
129 0.0 0.0 3.6
130 75.9 55.2 82.8





#29

Tetrahydrofuran

Concen: 58.618 ug/l

RT: 7.556 min Scan# 91

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC010





Tgt Ion: 42 Resp: 14275

Ion Ratio Lower Upper

42 100

72 39.9 34.6 52.0

71 41.5 33.0 49.6



#30

Chloroform

Concen: 12.341 ug/l

RT: 7.697 min Scan# 982

Delta R.T. -0.006 min

Lab File: VD073797.D




Acq: 06 Jul 2022 17:47


Tgt Ion: 83 Resp: 29895

Ion Ratio Lower Upper

83 100

85 59.1 53.4 80.0









#31
Cyclohexane
Concen: 11.974 ug/l
RT: 7.979 min Scan# 1029
Delta R.T. 0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 56 Resp: 20891
Ion Ratio Lower Upper
56 100
69 53.3 29.4 44.2#
84 88.9 72.1 108.1



#32
1,1,1-Trichloroethane
Concen: 11.525 ug/l
RT: 7.891 min Scan# 1015
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 97 Resp: 24313
Ion Ratio Lower Upper
97 100
99 67.3 49.8 74.8
61 50.8 35.8 53.6









#33
 1,2-Dichloroethane-d4
 Concen: 14.168 ug/l
 RT: 8.326 min Scan# 1088
 Delta R.T. 0.006 min
 Lab File: VD073797.D
 Acq: 06 Jul 2022 17:47




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDIC010


Tgt Ion: 65 Resp: 16016
 Ion Ratio Lower Upper
 65 100
 67 52.4 0.0 105.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.855 min Scan# 1179
 Delta R.T. 0.000 min
 Lab File: VD073797.D
 Acq: 06 Jul 2022 17:47

Tgt Ion: 114 Resp: 218581
 Ion Ratio Lower Upper
 114 100
 63 21.7 0.0 44.2
 88 16.9 0.0 33.6





#35

Dibromofluoromethane

Concen: 12.574 ug/l

RT: 7.908 min Scan# 1018

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC010





Tgt Ion:113 Resp: 15118

Ion Ratio Lower Upper

113 100

111 96.0 82.7 124.1

192 14.8 11.7 17.5



#36

1,1-Dichloropropene

Concen: 11.467 ug/l

RT: 8.103 min Scan# 1051

Delta R.T. -0.006 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47




Tgt Ion: 75 Resp: 18972


Ion Ratio Lower Upper

75 100

110 32.0 16.9 50.7

77 32.3 24.8 37.2





#37

Ethyl Acetate

Concen: 11.845 ug/l

RT: 7.285 min Scan# 913

Delta R.T. -0.006 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC010





Tgt Ion: 43 Resp: 10569

Ion Ratio Lower Upper

43 100

61 11.5 10.7 16.1

70 3.9 7.7 11.5#



#38

Carbon Tetrachloride

Concen: 11.141 ug/l

RT: 8.091 min Scan# 1049

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47




Tgt Ion: 117 Resp: 20032


Ion Ratio Lower Upper

117 100

119 88.1 79.3 118.9

121 27.5 22.9 34.3





#39

Methylcyclohexane

Concen: 10.316 ug/l

RT: 9.344 min Scan# 11

Delta R.T. -0.006 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :

VSTDICC010





Tgt Ion: 83 Resp: 19148

Ion Ratio Lower Upper

83 100

55 72.4 59.3 88.9

98 49.9 40.4 60.6



#40

Benzene

Concen: 11.051 ug/l

RT: 8.344 min Scan# 1092

Delta R.T. 0.000 min

Lab File: VD073797.D




Acq: 06 Jul 2022 17:47


Tgt Ion: 78 Resp: 54359

Ion Ratio Lower Upper

78 100

77 25.4 17.8 26.6









#41
Methacrylonitrile
Concen: 9.185 ug/l
RT: 7.508 min Scan# 91
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:	41	Resp:	5148
Ion	Ratio	Lower	Upper
41	100		
39	74.2	51.3	76.9
67	84.3	65.5	98.3
52	35.3	29.2	43.8



#42
1,2-Dichloroethane
Concen: 12.366 ug/l
RT: 8.420 min Scan# 1105
Delta R.T. 0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:	62	Resp:	18809
Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	17.4









#43
Isopropyl Acetate
Concen: 11.245 ug/l
RT: 8.450 min Scan# 1110
Delta R.T. 0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC010


Tgt Ion: 43 Resp: 21384
Ion Ratio Lower Upper
43 100
61 28.8 21.8 32.8
87 12.4 9.6 14.4



#44
Trichloroethene
Concen: 10.901 ug/l
RT: 9.102 min Scan# 1221
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:130 Resp: 14113
Ion Ratio Lower Upper
130 100
95 124.4 0.0 212.6





#45

1,2-Dichloropropane

Concen: 10.626 ug/l

RT: 9.379 min Scan# 1

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :





VSTDIC010

Tgt Ion: 63 Resp: 14994

Ion Ratio Lower Upper

63 100

65 27.9 23.4 35.2



#46

Dibromomethane

Concen: 11.801 ug/l

RT: 9.467 min Scan# 1283

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47




Tgt Ion: 93 Resp: 8811


Ion Ratio Lower Upper

93 100

95 80.1 64.8 97.2

174 74.8 59.3 88.9









#47
Bromodichloromethane
Concen: 11.508 ug/l
RT: 9.650 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC010


Tgt Ion:	83	Resp:	23273
Ion	Ratio	Lower	Upper
83	100		
85	63.9	49.4	74.0
127	8.4	7.8	11.6



#48
Methyl methacrylate
Concen: 11.522 ug/l
RT: 9.444 min Scan# 1279
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:	41	Resp:	9797
Ion	Ratio	Lower	Upper
41	100		
69	84.4	64.6	96.8
39	71.3	54.2	81.4





#49

1,4-Dioxane

Concen: 209.013 ug/l

RT: 9.455 min Scan# 1281

Delta R.T. 0.006 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC010





Tgt Ion: 88 Resp: 2102

Ion Ratio Lower Upper

88 100

43 26.4 33.3 49.9#

58 59.8 54.2 81.4



#50

Toluene-d8

Concen: 12.232 ug/l

RT: 10.326 min Scan# 1429

Delta R.T. -0.006 min

Lab File: VD073797.D




Acq: 06 Jul 2022 17:47


Tgt Ion: 98 Resp: 53454

Ion Ratio Lower Upper

98 100

100 63.8 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 57.697 ug/l
RT: 10.220 min Scan# 1411
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC010


Tgt Ion: 43 Resp: 56228
Ion Ratio Lower Upper
43 100
58 34.9 30.1 45.1



#52
Toluene
Concen: 10.758 ug/l
RT: 10.397 min Scan# 1441
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 92 Resp: 35303
Ion Ratio Lower Upper
92 100
91 172.2 137.4 206.0









#53
t-1,3-Dichloropropene
Concen: 10.746 ug/l
RT: 10.608 min Scan# 1477
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 75 Resp: 21214
Ion Ratio Lower Upper
75 100
77 32.3 26.6 40.0



#54
cis-1,3-Dichloropropene
Concen: 10.998 ug/l
RT: 10.079 min Scan# 1387
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 75 Resp: 24424
Ion Ratio Lower Upper
75 100
77 33.4 25.4 38.0
39 57.6 42.9 64.3









#55
1,1,2-Trichloroethane
Concen: 10.402 ug/l
RT: 10.791 min Scan# 1508
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 97	Resp: 11546
Ion Ratio	Lower Upper
97 100	
83 77.6	68.2 102.2
85 56.9	43.9 65.9
99 71.8	47.8 71.8#



#56
Ethyl methacrylate
Concen: 10.121 ug/l
RT: 10.649 min Scan# 1484
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 69	Resp: 14837
Ion Ratio	Lower Upper
69 100	
41 77.7	58.1 87.1
39 39.5	36.6 54.8









#57
1,3-Dichloropropane
Concen: 11.288 ug/l
RT: 10.932 min Scan# 1533
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 76 Resp: 21134
Ion Ratio Lower Upper
76 100
78 32.2 26.6 40.0



#58
2-Chloroethyl Vinyl ether
Concen: 50.464 ug/l
RT: 9.932 min Scan# 1362
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 63 Resp: 29592
Ion Ratio Lower Upper
63 100
106 28.0 22.2 33.4









#59
2-Hexanone
Concen: 53.541 ug/l
RT: 10.973 min Scan# 1539
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 35427
Ion Ratio Lower Upper
43 100
58 55.2 26.2 78.6



#60
Dibromochloromethane
Concen: 10.755 ug/l
RT: 11.126 min Scan# 1565
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 129 Resp: 14567
Ion Ratio Lower Upper
129 100
127 77.5 38.5 115.5









#61
1,2-Dibromoethane
Concen: 11.148 ug/l
RT: 11.238 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:107 Resp: 11373
Ion Ratio Lower Upper
107 100
109 88.4 78.6 117.8



#62
4-Bromofluorobenzene
Concen: 11.889 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 95 Resp: 20638
Ion Ratio Lower Upper
95 100
174 61.8 0.0 132.4
176 62.6 0.0 129.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.632 min Scan# 1651
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:	117	Resp:	199904
Ion	Ratio	Lower	Upper
117	100		
82	59.3	48.2	72.2
119	30.4	25.2	37.8



#64
Tetrachloroethene
Concen: 11.015 ug/l
RT: 10.867 min Scan# 1521
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:	164	Resp:	11235
Ion	Ratio	Lower	Upper
164	100		
166	115.2	98.8	148.2
129	99.8	91.0	136.4
131	90.4	84.5	126.7









#65
Chlorobenzene
Concen: 10.958 ug/l
RT: 11.661 min Scan# 1656
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:112 Resp: 40685
Ion Ratio Lower Upper
112 100
114 29.4 26.6 39.8



#66
1,1,1,2-Tetrachloroethane
Concen: 10.976 ug/l
RT: 11.732 min Scan# 1668
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:131 Resp: 15419
Ion Ratio Lower Upper
131 100
133 94.4 48.4 145.3
119 66.9 33.1 99.3





#67

Ethyl Benzene

Concen: 10.696 ug/l

RT: 11.732 min Scan# 1668

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :





VSTDICC010

Tgt Ion: 91 Resp: 72638

Ion Ratio Lower Upper

91 100

106 32.0 24.1 36.1



#68

m/p-Xylenes

Concen: 20.820 ug/l

RT: 11.838 min Scan# 1686

Delta R.T. -0.006 min

Lab File: VD073797.D




Acq: 06 Jul 2022 17:47


Tgt Ion: 106 Resp: 53881

Ion Ratio Lower Upper

106 100

91 207.9 165.4 248.2









#69
o-Xylene
Concen: 10.486 ug/l
RT: 12.173 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC010


Tgt Ion:106 Resp: 26501
Ion Ratio Lower Upper
106 100
91 212.8 112.3 336.8



#70
Styrene
Concen: 10.346 ug/l
RT: 12.179 min Scan# 1744
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:104 Resp: 45040
Ion Ratio Lower Upper
104 100
78 52.2 42.2 63.2
103 52.0 43.9 65.9









#71
Bromoform
Concen: 10.276 ug/l
RT: 12.344 min Scan# 11
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:173 Resp: 7614
Ion Ratio Lower Upper
173 100
175 45.9 23.8 71.5
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:152 Resp: 89174
Ion Ratio Lower Upper
152 100
115 65.9 32.2 96.6
150 165.8 0.0 353.0









#73
Isopropylbenzene
Concen: 10.444 ug/l
RT: 12.467 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:105 Resp: 70822
Ion Ratio Lower Upper
105 100
120 27.3 12.9 38.7



#74
N-ethyl acetate
Concen: 10.648 ug/l
RT: 12.273 min Scan# 1760
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 43 Resp: 19824
Ion Ratio Lower Upper
43 100
70 39.1 31.5 47.3
55 23.7 18.5 27.7
61 21.7 18.4 27.6









#75
1,1,2,2-Tetrachloroethane
Concen: 11.084 ug/l
RT: 12.714 min Scan# 1835
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 83 Resp: 15165
Ion Ratio Lower Upper
83 100
131 9.5 5.2 15.6
85 64.4 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 21.592 ug/l
RT: 12.773 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 75 Resp: 18300
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 11.066 ug/l

RT: 12.743 min Scan# 1840

Delta R.T. -0.006 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC010





Tgt Ion:156 Resp: 15359

Ion Ratio Lower Upper

156 100

77 216.1 110.8 332.4

158 96.3 49.1 147.3



#78

n-propylbenzene

Concen: 10.622 ug/l

RT: 12.808 min Scan# 1851

Delta R.T. 0.000 min

Lab File: VD073797.D




Acq: 06 Jul 2022 17:47


Tgt Ion: 91 Resp: 90001

Ion Ratio Lower Upper

91 100

120 21.7 10.9 32.7









#79
2-Chlorotoluene
Concen: 10.589 ug/l
RT: 12.896 min Scan# 1866
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 91 Resp: 50349
Ion Ratio Lower Upper
91 100
126 32.9 16.1 48.2



#80
1,3,5-Trimethylbenzene
Concen: 10.749 ug/l
RT: 12.943 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 105 Resp: 60759
Ion Ratio Lower Upper
105 100
120 45.2 23.8 71.4









#81
trans-1,4-Dichloro-2-butene
Concen: 8.996 ug/l
RT: 12.520 min Scan# 1801
Delta R.T. 0.006 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion: 75 Resp: 3970
Ion Ratio Lower Upper
75 100
53 103.0 77.6 116.4
89 48.3 37.2 55.8



#82
4-Chlorotoluene
Concen: 10.615 ug/l
RT: 12.991 min Scan# 1882
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 91 Resp: 52555
Ion Ratio Lower Upper
91 100
126 33.5 16.1 48.2









#83
tert-Butylbenzene
Concen: 10.497 ug/l
RT: 13.208 min Scan# 1919
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:119 Resp: 52691
Ion Ratio Lower Upper
119 100
91 69.8 33.7 101.1
134 23.3 12.4 37.0



#84
1,2,4-Trimethylbenzene
Concen: 10.508 ug/l
RT: 13.255 min Scan# 1927
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:105 Resp: 58596
Ion Ratio Lower Upper
105 100
120 46.0 22.1 66.5





#85

sec-Butylbenzene

Concen: 10.464 ug/l

RT: 13.385 min Scan# 1949

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :





VSTDIC010

Tgt Ion:105 Resp: 80616

Ion Ratio Lower Upper

105 100

134 19.2 9.4 28.2



#86

p-Isopropyltoluene

Concen: 10.368 ug/l

RT: 13.502 min Scan# 1969

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47




Tgt Ion:119 Resp: 64018


Ion Ratio Lower Upper

119 100

134 24.0 12.7 38.1

91 26.5 13.2 39.6





#87

1,3-Dichlorobenzene

Concen: 10.684 ug/l

RT: 13.502 min Scan# 1969

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC010





Tgt Ion:146 Resp: 31060

Ion Ratio Lower Upper

146 100

111 44.4 21.6 65.0

148 60.7 31.5 94.5



#88

1,4-Dichlorobenzene

Concen: 11.034 ug/l

RT: 13.579 min Scan# 1982

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47




Tgt Ion:146 Resp: 31971


Ion Ratio Lower Upper

146 100

111 48.8 21.6 64.6

148 62.8 31.4 94.0









#89
n-Butylbenzene
Concen: 10.599 ug/l
RT: 13.826 min Scan# 2024
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDIC010


Tgt Ion:	91	Resp:	66028
Ion	Ratio	Lower	Upper
91	100		
92	52.6	26.9	80.6
134	23.6	11.9	35.9



#90
Hexachloroethane
Concen: 10.323 ug/l
RT: 14.096 min Scan# 2070
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion:	117	Resp:	12743
Ion	Ratio	Lower	Upper
117	100		
201	54.1	26.6	79.8









#91
1,2-Dichlorobenzene
Concen: 10.542 ug/l
RT: 13.873 min Scan# 2032
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47




Instrument :
MSVOA_D
ClientSampleId :
VSTDICC010


Tgt Ion:146 Resp: 27255
Ion Ratio Lower Upper
146 100
111 48.7 22.3 66.9
148 62.1 31.4 94.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 10.443 ug/l
RT: 14.490 min Scan# 2137
Delta R.T. 0.000 min
Lab File: VD073797.D
Acq: 06 Jul 2022 17:47

Tgt Ion: 75 Resp: 2390
Ion Ratio Lower Upper
75 100
155 74.9 36.4 109.3
157 96.9 43.5 130.4





#93

1,2,4-Trichlorobenzene

Concen: 10.581 ug/l

RT: 15.143 min Scan# 2248

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC010





Tgt Ion:180 Resp: 16098

Ion Ratio Lower Upper

180 100

182 93.4 46.9 140.6

145 36.0 17.5 52.5



#94

Hexachlorobutadiene

Concen: 10.173 ug/l

RT: 15.249 min Scan# 2266

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47




Tgt Ion:225 Resp: 7642


Ion Ratio Lower Upper

225 100

223 66.2 31.5 94.5

227 66.9 31.6 94.7





#95

Naphthalene

Concen: 10.310 ug/l

RT: 15.379 min Scan# 21

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Instrument :

MSVOA_D

ClientSampleId :

VSTDIC010





Tgt Ion:128 Resp: 35035

Ion Ratio Lower Upper

128 100

127 13.2 10.4 15.6

129 11.7 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 10.379 ug/l

RT: 15.573 min Scan# 2321

Delta R.T. 0.000 min

Lab File: VD073797.D

Acq: 06 Jul 2022 17:47

Tgt Ion:180 Resp: 13815

Ion Ratio Lower Upper

180 100

182 97.9 46.0 138.0

145 37.0 18.1 54.4

