



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070622\
 Data File : VD073787.D
 Acq On : 06 Jul 2022 09:32
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
 Title : SW846 8260
 Last Update : Fri Jul 08 05:39:57 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	28000	PASS
75	95	30	60	56.8	70536	PASS
95	95	100	100	100.0	124219	PASS
96	95	5	9	6.9	8585	PASS
173	174	0.00	2	1.6	1296	PASS
174	95	50	100	65.6	81525	PASS
175	174	5	9	8.6	6989	PASS
176	174	95	101	96.7	78851	PASS
177	176	5	9	7.5	5930	PASS