

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070722\
 Data File : VD073814.D
 Acq On : 07 Jul 2022 17:58
 Operator : VA/SY
 Sample : N3593-05
 Misc : 8.40G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20220484B-2B

Quant Time: Jul 08 06:34:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 05:39:57 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	98299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	219858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	214041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	97716	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	83369	68.931	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 137.860%	
35) Dibromofluoromethane	7.909	113	73728	54.653	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.300%	
50) Toluene-d8	10.332	98	280880	61.462	ug/l	0.00
Spiked Amount	50.000	Range	78 - 12	Recovery	= 122.920%#	
62) 4-Bromofluorobenzene	12.620	95	110445	58.796	ug/l	0.00
Spiked Amount	50.000	Range	50 - 14	Recovery	= 117.600%#	
Target Compounds						Qvalue
16) Acetone	4.156	43	8738	19.967	ug/l	95
17) Carbon Disulfide	4.403	76	103304	101.980	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070722\
Data File : VD073814.D
Acq On : 07 Jul 2022 17:58
Operator : VA/SY
Sample : N3593-05
Misc : 8.40G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20220484B-2B

Quant Time: Jul 08 06:34:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 05:39:57 2022
Response via : Initial Calibration

