

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070722\
 Data File : VD073800.D
 Acq On : 07 Jul 2022 09:49
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 08 06:28:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 05:39:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.335	0.261	22.1	106	0.00
3 P	Chloromethane	0.446	0.347	22.2	97	0.00
4 C	Vinyl Chloride	0.537	0.446	16.9#	105	0.00
5 T	Bromomethane	0.418	0.325	22.2	102	0.00
6 T	Chloroethane	0.409	0.346	15.4	102	0.00
7 T	Trichlorofluoromethane	0.836	0.722	13.6	103	0.00
8 T	Diethyl Ether	0.256	0.243	5.1	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.517	0.470	9.1	110	0.00
10 T	Methyl Iodide	0.340	0.324	4.7	110	0.00
11 T	Tert butyl alcohol	0.210	0.090	57.1#	93	-0.03
12 CM	1,1-Dichloroethene	0.410	0.344	16.1#	98	0.00
13 T	Acrolein	0.035	0.042	-20.0	87	0.00
14 T	Allyl chloride	0.774	0.692	10.6	101	0.00
15 T	Acrylonitrile	0.147	0.140	4.8	99	0.00
16 T	Acetone	0.132	0.127	3.8	119	0.00
17 T	Carbon Disulfide	0.668	0.504	24.6	98	-0.01
18 T	Methyl Acetate	0.435	0.301	30.8#	97	0.00
19 T	Methyl tert-butyl Ether	1.511	1.423	5.8	104	0.00
20 T	Methylene Chloride	1.143	0.539	52.8#	100	0.00
21 T	trans-1,2-Dichloroethene	0.488	0.411	15.8	101	0.00
22 T	Diisopropyl ether	1.997	1.935	3.1	104	0.00
23 T	Vinyl Acetate	1.070	1.094	-2.2	101	0.00
24 P	1,1-Dichloroethane	1.122	1.019	9.2	100	0.00
25 T	2-Butanone	0.202	0.192	5.0	101	0.00
26 T	2,2-Dichloropropane	1.045	1.015	2.9	114	0.00
27 T	cis-1,2-Dichloroethene	0.687	0.613	10.8	100	0.00
28 T	Bromochloromethane	0.437	0.432	1.1	103	-0.01
29 T	Tetrahydrofuran	0.119	0.115	3.4	102	0.00
30 C	Chloroform	1.244	1.190	4.3#	105	0.00
31 T	Cyclohexane	0.794	0.606	23.7	101	0.00
32 T	1,1,1-Trichloroethane	1.053	0.990	6.0	104	0.00
33 S	1,2-Dichloroethane-d4	0.615	0.510	17.1	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.307	0.280	8.8	108	0.00
36 T	1,1-Dichloropropene	0.393	0.351	10.7	104	0.00
37 T	Ethyl Acetate	0.234	0.218	6.8	100	0.00
38 T	Carbon Tetrachloride	0.446	0.418	6.3	105	0.00
39 T	Methylcyclohexane	0.405	0.349	13.8	103	0.00
40 TM	Benzene	1.160	1.056	9.0	101	0.00
41 T	Methacrylonitrile	0.129	0.139	-7.8	121	0.00
42 TM	1,2-Dichloroethane	0.399	0.377	5.5	100	0.00
43 T	Isopropyl Acetate	0.472	0.448	5.1	99	0.00
44 TM	Trichloroethene	0.309	0.274	11.3	103	0.00
45 C	1,2-Dichloropropane	0.333	0.317	4.8#	100	0.00
46 T	Dibromomethane	0.191	0.175	8.4	99	0.00
47 T	Bromodichloromethane	0.517	0.504	2.5	104	0.00
48 T	Methyl methacrylate	0.215	0.212	1.4	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070722\
 Data File : VD073800.D
 Acq On : 07 Jul 2022 09:49
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 08 06:28:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 05:39:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	99	0.00
50 S	Toluene-d8	1.039	0.867	16.6	105	0.00
51 T	4-Methyl-2-Pentanone	0.246	0.239	2.8	99	0.00
52 CM	Toluene	0.768	0.710	7.6#	101	0.00
53 T	t-1,3-Dichloropropene	0.478	0.475	0.6	105	0.00
54 T	cis-1,3-Dichloropropene	0.530	0.506	4.5	103	0.00
55 T	1,1,2-Trichloroethane	0.268	0.266	0.7	103	0.00
56 T	Ethyl methacrylate	0.342	0.338	1.2	101	0.00
57 T	1,3-Dichloropropane	0.460	0.437	5.0	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.135	0.120	11.1	97	0.00
59 T	2-Hexanone	0.164	0.163	0.6	101	0.00
60 T	Dibromochloromethane	0.330	0.334	-1.2	104	0.00
61 T	1,2-Dibromoethane	0.246	0.230	6.5	99	0.00
62 S	4-Bromofluorobenzene	0.427	0.380	11.0	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.252	0.231	8.3	109	0.00
65 PM	Chlorobenzene	0.955	0.905	5.2	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.371	-0.8	106	0.00
67 C	Ethyl Benzene	1.732	1.651	4.7#	105	0.00
68 T	m/p-Xylenes	0.655	0.614	6.3	103	0.00
69 T	o-Xylene	0.631	0.610	3.3	104	0.00
70 T	Styrene	1.107	1.058	4.4	101	0.00
71 P	Bromoform	0.189	0.192	-1.6	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.849	3.705	3.7	103	0.00
74 T	N-amyl acetate	1.086	1.070	1.5	101	0.00
75 P	1,1,2,2-Tetrachloroethane	0.800	0.784	2.0	100	0.00
76 T	1,2,3-Trichloropropane	0.544	0.639	-17.5	143	0.00
77 T	Bromobenzene	0.801	0.773	3.5	106	0.00
78 T	n-propylbenzene	4.824	4.598	4.7	105	0.00
79 T	2-Chlorotoluene	2.729	2.615	4.2	102	0.00
80 T	1,3,5-Trimethylbenzene	3.216	3.107	3.4	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.244	0.248	-1.6	103	0.00
82 T	4-Chlorotoluene	2.877	2.783	3.3	103	0.00
83 T	tert-Butylbenzene	2.849	2.742	3.8	101	0.00
84 T	1,2,4-Trimethylbenzene	3.184	3.078	3.3	105	0.00
85 T	sec-Butylbenzene	4.362	4.192	3.9	105	0.00
86 T	p-Isopropyltoluene	3.478	3.412	1.9	106	0.00
87 T	1,3-Dichlorobenzene	1.681	1.605	4.5	102	0.00
88 T	1,4-Dichlorobenzene	1.701	1.599	6.0	102	0.00
89 T	n-Butylbenzene	3.514	3.403	3.2	105	0.00
90 T	Hexachloroethane	0.701	0.659	6.0	100	0.00
91 T	1,2-Dichlorobenzene	1.488	1.442	3.1	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.133	0.126	5.3	96	0.00
93 T	1,2,4-Trichlorobenzene	0.851	0.842	1.1	106	0.00
94 T	Hexachlorobutadiene	0.410	0.390	4.9	105	0.00
95 T	Naphthalene	1.922	1.881	2.1	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070722\
Data File : VD073800.D
Acq On : 07 Jul 2022 09:49
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jul 08 06:28:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D070622S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 05:39:57 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.752	0.720	4.3	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6