

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070821\
 Data File : VD069628.D
 Acq On : 08 Jul 2021 14:41
 Operator : VA/SY
 Sample : M2967-04
 Misc : 3.83G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 GRACE-AVE-5-FEET

Quant Time: Jul 09 01:02:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	591805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1155927	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1251437	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	567342	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	360954	48.931	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.860%
35) Dibromofluoromethane	7.914	113	381320	48.801	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.600%
50) Toluene-d8	10.338	98	1513047	51.624	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.240%
62) 4-Bromofluorobenzene	12.626	95	578274	60.015	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	120.020%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.956	84	101723	7.440	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070821\
Data File : VD069628.D
Acq On : 08 Jul 2021 14:41
Operator : VA/SY
Sample : M2967-04
Misc : 3.83G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
GRACE-AVE-5-FEET

Quant Time: Jul 09 01:02:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

