

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070821\
 Data File : VD069636.D
 Acq On : 08 Jul 2021 18:29
 Operator : VA/SY
 Sample : M2969-11
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB06

Quant Time: Jul 09 01:05:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	558204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1087470	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	1197386	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	536689	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	348235	50.049	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.100%
35) Dibromofluoromethane	7.914	113	350477	47.677	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.360%
50) Toluene-d8	10.338	98	1434867	52.038	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.080%
62) 4-Bromofluorobenzene	12.626	95	548816	60.543	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	121.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070821\
Data File : VD069636.D
Acq On : 08 Jul 2021 18:29
Operator : VA/SY
Sample : M2969-11
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB06

Quant Time: Jul 09 01:05:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

