

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070821\
 Data File : VD069637.D
 Acq On : 08 Jul 2021 18:57
 Operator : VA/SY
 Sample : M2969-13
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB07

Quant Time: Jul 09 01:05:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	555302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1086393	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1161059	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	505361	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	336778	48.655	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	97.320%		
35) Dibromofluoromethane	7.914	113	272826	37.150	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	74.300%		
50) Toluene-d8	10.338	98	1404644	50.992	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	101.980%		
62) 4-Bromofluorobenzene	12.626	95	524894	57.961	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	115.920%		
Target Compounds						Qvalue
16) Acetone	4.162	43	13064	11.213	ug/l	93
20) Methylene Chloride	4.968	84	89921	6.586	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070821\
Data File : VD069637.D
Acq On : 08 Jul 2021 18:57
Operator : VA/SY
Sample : M2969-13
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB07

Quant Time: Jul 09 01:05:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

